

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-004935 Location: TOXL

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 784 psi Regulator: 750 psi
754 CEE

B. Tests (Sign and attach test records):

- Configure simulator for the following tests (Level 1,S).
1. Wet Calibration Check
 - a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: MP3002 Lot #: 17160 Exp. Date: 6.8.19
 - b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.400 AC
Sim. Ser #: MP3061 Lot #: 201605F Exp. Date: 5.19.18

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR7089 Lot #: ICS #5 Exp. Date: 5.15.19
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 19817080A4 Cyl #: 21 Exp. Date: 9.5.19

Test 1	<u>0.078</u>	Test 4	<u>0.079</u>	Test 7	<u>0.079</u>
Test 2	<u>0.079</u>	Test 5	<u>0.078</u>	Test 8	<u>0.079</u>
Test 3	<u>0.078</u>	Test 6	<u>0.079</u>	Test 9	<u>0.078</u>
Average	<u>0.079</u>				

C. Remarks/Maintenance: ANNUAL INSPECTION, REPLACED SIMULATOR
RETURN AND BREATH HOSE O-RINGS, TANK SENSOR CALIBRATION

Instrument is acceptable to be used in the field.

Charles E. Ede
Breath Analyst Signature

4.10.2018
Date

Roberta Mueger-Nimmer
Reviewed by

4.17.2018
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004935
Location = TOXL 8164.14.00 09/16
04/10/2018 15:09

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? No
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 19817080A4
Standard Cyl #? 21
Standard Expiration? 09/05/2019
Oper No? 666666

Flow Cal. Date: 07/12/2012
Slope 655
Intercept -510289

IR Calibration Date: 05/20/2015

	3um	9um
0th Coef(*100):	-7871	-19684
1st Coef(*100):	264505	133108
2nd Coef(*100):	2865	1320
H2O adj(mg/l*10k):	200	298

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks: PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004935
Location = TOXL 8164.14.00 09/16
04/10/2018 15:14

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:15
02 Std. Sol.	0.018	15:16
03 Room Air	0.000	15:16
04 Std. Sol.	0.018	15:17
05 Room Air	0.000	15:17
06 Std. Sol.	0.019	15:18
07 Room Air	0.000	15:19

08 Sim Temp = 34.0°C

Simul Ser No = MP3002

Std Sol No = 17160

County = 08

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Low AC

0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004935
Location = TOXL 8164.14.00 09/16
04/10/2018 15:20

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:21
02 Std. Sol.	0.400	15:22
03 Room Air	0.000	15:22
04 Std. Sol.	0.404	15:23
05 Room Air	0.000	15:24
06 Std. Sol.	0.402	15:24
07 Room Air	0.000	15:25

08 Sim Temp = 34.0°C

Simul Ser No = MP3061
Std Sol No = 201605F
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

HIGH AC
0.400 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004935
Location = TOXL 8164.14.00 09/16
04/10/2018 15:26

Test	AC	Time
01 Room Air	0.000	15:27
02 *Subject Test	INT*	15:27
03 Room Air	0.000	15:28

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 19817080A4
Cyl No = 21
Expiration Date = 09/05/2019
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Interferent Check*
0.10 Ac + 0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004935
Location = TOXL 8164.14.00 09/16
04/10/2018 15:28

Test	AC	Time
01 Diagnostic	OK	15:29
02 Room Air	0.000	15:29
03 *Subject Test	RFI*	15:29
04 Room Air	0.000	15:30

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 19817080A4
Cyl No = 21
Expiration Date = 09/05/2019
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004935
Location = TOXL 8164.14.00 09/16
04/10/2018 15:30

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:30
02 Std. Gas	0.078	15:31
03 Room Air	0.000	15:31
04 Std. Gas	0.079	15:32
05 Room Air	0.000	15:32
06 Std. Gas	0.078	15:33
07 Room Air	0.000	15:33

Lot No = 19817080A4

Cyl No = 21

Exp Date = 09/05/2019

County = 08

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

CALIBRATION CHECK

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004935
Location = TOXL 8164.14.00 09/16
04/10/2018 15:33

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:34
02 Std. Gas	0.079	15:34
03 Room Air	0.000	15:35
04 Std. Gas	0.078	15:35
05 Room Air	0.000	15:35
06 Std. Gas	0.079	15:36
07 Room Air	0.000	15:36


Lot No = 19817080A4

Cyl No = 21

Exp Date = 09/05/2019

County = 08

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

CALIBRATION CHECK
0.080 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004935
Location = TOXL 8164.14.00 09/16
04/10/2018 15:37

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:37
02 Std. Gas	0.079	15:37
03 Room Air	0.000	15:38
04 Std. Gas	0.079	15:38
05 Room Air	0.000	15:39
06 Std. Gas	0.078	15:39
07 Room Air	0.000	15:40

Lot No = 19817080A4
Cyl No = 21
Exp Date = 09/05/2019
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0-080 AC