

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-004947

Location: ToxV

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 730 psi Regulator: 740 psi

B. Tests (Sign and attach test records):

- Configure simulator for the following tests (Level 1,S).
1. Wet Calibration Check
 - a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: MP3002 Lot #: 17160 Exp. Date: 6-8-19
 - b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.400 AC
Sim. Ser #: MP3061 Lot #: 201605F Exp. Date: 5-19-18

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR7089 Lot #: ICS#5 Exp. Date: 5-15-19
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 19817080A4 Cyl #: 21 Exp. Date: 9-5-19

Test 1	<u>0.082</u>	Test 4	<u>0.082</u>	Test 7	<u>0.082</u>
Test 2	<u>0.081</u>	Test 5	<u>0.083</u>	Test 8	<u>0.082</u>
Test 3	<u>0.083</u>	Test 6	<u>0.082</u>	Test 9	<u>0.082</u>
Average <u>0.082</u>					

C. Remarks/Maintenance: ANNUAL INSPECTION

Instrument is acceptable to be used in the field.

Charles E. Eda
Breath Analyst Signature

4.10.18
Date

Roseta Muegu-romo
Reviewed by

4.17.18
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004947
Location = TOXL 8164.14.00 09/16
04/10/2018 15:50

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 19817080A4
Standard Cyl #? 21
Standard Expiration? 09/05/2019
Oper No? 666666

Flow Cal. Date: 01/19/2010
Slope 679
Intercept -681178

IR Calibration Date: 05/12/2015
 3um 9um

	3um	9um
0th Coef(*100):	-14993	-20474
1st Coef(*100):	257346	132154
2nd Coef(*100):	2870	1330
H2O adj(mg/l*10k):	492	531

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004947
Location = TOXL 8164.14.00 09/16
04/10/2018 15:53

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:54
02 Std. Sol.	0.018	15:55
03 Room Air	0.000	15:55
04 Std. Sol.	0.018	15:56
05 Room Air	0.000	15:56
06 Std. Sol.	0.019	15:57
07 Room Air	0.000	15:57

08 Sim Temp = 34.0°C

Simul Ser No = MP3002

Std Sol No = 17160

County = 08

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

Low AC
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004947
Location = TOXL 8164.14.00 09/16
04/10/2018 16:03

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:03
02 Std. Sol.	0.391	16:04
03 Room Air	0.000	16:05
04 Std. Sol.	0.391	16:05
05 Room Air	0.000	16:06
06 Std. Sol.	0.392	16:07
07 Room Air	0.000	16:07

08 Sim Temp = 34.0°C

Simul Ser No = MP3061

Std Sol No = 201605F

County = 08

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

HIGH AC
0.400 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004947
Location = TOXL 8164.14.00 09/16
04/10/2018 16:08

Test	AC	Time
01 Room Air	0.000	16:08
02 *Subject Test	INT*	16:09
03 Room Air	0.000	16:10

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 19817080A4
Cyl No = 21
Expiration Date = 09/05/2019
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Interferent Check*
0.10 AC + 0.05% ACETONE

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004947
Location = TOXL 8164.14.00 09/16
04/10/2018 16:10

Test	AC	Time
01 Diagnostic	OK	16:11
02 Room Air	0.000	16:11
03 Reference	RFI*	16:11
04 Room Air	0.000	16:12

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 19817080A4
Cyl No = 21
Expiration Date = 09/05/2019
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004947
Location = TOXL 8164.14.00 09/16
04/10/2018 16:12

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:13
02 Std. Gas	0.082	16:13
03 Room Air	0.000	16:14
04 Std. Gas	0.081	16:14
05 Room Air	0.000	16:14
06 Std. Gas	0.083	16:15
07 Room Air	0.000	16:15

Lot No = 19817080A4

Cyl No = 21

Exp Date = 09/05/2019

County = 08

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION CHECK

0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004947
Location = TOXL 8164.14.00 09/16
04/10/2018 16:15

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:16
02 Std. Gas	0.082	16:16
03 Room Air	0.000	16:17
04 Std. Gas	0.083	16:17
05 Room Air	0.000	16:18
06 Std. Gas	0.082	16:18
07 Room Air	0.000	16:19

Lot No = 19817080A4
Cyl No = 21
Exp Date = 09/05/2019
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION CHECK
0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004947
Location = TOXL 8164.14.00 09/16
04/10/2018 16:19

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:19
02 Std. Gas	0.082	16:20
03 Room Air	0.000	16:20
04 Std. Gas	0.082	16:20
05 Room Air	0.000	16:21
06 Std. Gas	0.082	16:21
07 Room Air	0.000	16:22

Lot No = 19817080A4
Cyl No = 21
Exp Date = 09/05/2019
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000