

**INTOXILYZER® 8000 INSPECTION**

Intoxilyzer® 8000 Serial Number: 80-004946

Location: TOXL

A. General Setup and Checks:

1.  Diagnostic Tests Pass and Instrument in Ready Mode
2.  Breath tube heated
3.  Date, time and location code (Level 2,E). Re-set if necessary.
4.  Print test (Level 1,P). Sign and attach test record.
5.  Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
  - a. Display: 723 psi Regulator: 725 psi

B. Tests (Sign and attach test records):

- Configure simulator for the following tests (Level 1,S).
1. Wet Calibration Check
    - a.  Low AC (Level 1,C): Known Value  $\leq 0.03$  AC: 0.020 AC  
Sim. Ser #: MP3002 Lot #: 17160 Exp. Date: 6.8.19
    - b.  High AC (Level 1,C): Known Value  $\geq 0.25$  AC: 0.300 AC  
Sim. Ser #: MP3065 Lot #: 17350 Exp. Date: 10.11.19

Configure dry gas standard for the following tests (Level 1,S).

2.  Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone  
Sim. Ser #: DR7089 Lot #: ICS#5 Exp. Date: 5.15.19  
 Display reads "Interferent Detect"
3.  RFI Check (CMS Mode)  Display reads "RFI Detect"
4.  Dry Calibration Check (Level 1,C): Known Value 0.080 AC  
Gas Cyl Lot #: 19817080A4 Cyl #: 21 Exp. Date: 9.5.19

Test 1	<u>0.080</u>	Test 4	<u>0.080</u>	Test 7	<u>0.080</u>
Test 2	<u>0.080</u>	Test 5	<u>0.080</u>	Test 8	<u>0.081</u>
Test 3	<u>0.080</u>	Test 6	<u>0.080</u>	Test 9	<u>0.080</u>
Average	<u>0.080</u>				

C. Remarks/Maintenance: ANNUAL INSPECTION, REPLACED SIMULATOR RETURN AND BREATH HOSE O-RINGS

Instrument is acceptable to be used in the field.

Charles E. Eden  
Breath Analyst Signature

4.12.18  
Date

Roberta Quezada-Nunno  
Reviewed by

4.18.18  
Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004946  
Location = TOXL                    8164.14.00 09/16  
04/12/2018                            14:12

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxy1234567890-=  
ABCDEFGHIJKLMN0PQRSTUVWXYZ!@#%\$^&\*()\_+?

abcdefghijklmnopqrstuvwxy1234567890-=  
ABCDEFGHIJKLMN0PQRSTUVWXYZ!@#%\$^&\*()\_+?

Current Instrument Setup


Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        No  
# of Print Copies?         1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            19817080A4  
Standard Cyl #?            21  
Standard Expiration?       09/05/2019  
Oper No?                    666666

Flow Cal. Date:            01/18/2010  
Slope                       717  
Intercept                  -728279

IR Calibration Date:        04/23/2015

	3um	9um
0th Coef(*100):	-7851	-17700
1st Coef(*100):	259859	133464
2nd Coef(*100):	3769	1675
H2O adj(mg/l*10k):	294	408

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
CHARLES EDER

Remarks:            PRINT TEST

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004946  
Location = TOXL      8164.14.00 09/16  
04/12/2018      14:13

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:14
02 Std. Sol.	0.018	14:15
03 Room Air	0.000	14:15
04 Std. Sol.	0.018	14:16
05 Room Air	0.000	14:16
06 Std. Sol.	0.017	14:17
07 Room Air	0.000	14:18

08 Sim Temp = 34.0°C

Simul Ser No = MP3002

Std Sol No = 17160

County = 08

Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

Low AC  
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004946  
Location = TOXL      8164.14.00 09/16  
04/12/2018      14:25

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:26
02 Std. Sol.	0.299	14:27
03 Room Air	0.000	14:27
04 Std. Sol.	0.302	14:28
05 Room Air	0.000	14:29
06 Std. Sol.	0.301	14:29
07 Room Air	0.000	14:30

08 Sim Temp = 34.0°C

Simul Ser No = MP3065  
Std Sol No = 17350  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: HIGH AC  
0.300 AC

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004946  
Location = TOXL      8164.14.00 09/16  
04/12/2018      14:31

Test	AC	Time
01 Room Air	0.000	14:31
02 *Subject Test	INT*	14:32
03 Room Air	0.000	14:33

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = INTERFERENT  
Dr. Lic. = ND/DIS821456  
Lot No = 19817080A4  
Cyl No = 21  
Expiration Date = 09/05/2019  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*Interferent Check*  
*0.10 AC + 0.05% ACETONE*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer                      Alcohol Analyzer  
North Dakota Model 8000                      SN 80-004946  
Location = TOXL                              8164.14.00 09/16  
04/12/2018                                              14:35

Test	AC	Time
01 Diagnostic	OK	14:36
02 Room Air	0.000	14:36
03 *Subject Test	RFI*	14:36
04 Room Air	0.000	14:37

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                              Weight = 150  
Test = DUI                                      Cit = RFI CHECK  
Dr. Lic. = ND/DIS821456  
Lot No = 19817080A4  
Cyl No = 21  
Expiration Date = 09/05/2019  
County = 08                                      Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks:

*RFI CHECK*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004946  
Location = TOXL      8164.14.00 09/16  
04/12/2018      14:38

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:38
02 Std. Gas	0.080	14:38
03 Room Air	0.000	14:39
04 Std. Gas	0.080	14:39
05 Room Air	0.000	14:40
06 Std. Gas	0.080	14:40
07 Room Air	0.000	14:41

Lot No = 19817080A4  
Cyl No = 21  
Exp Date = 09/05/2019  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080 AC

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004946  
Location = TOXL      8164.14.00 09/16  
04/12/2018      14:41

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:41
02 Std. Gas	0.080	14:42
03 Room Air	0.000	14:42
04 Std. Gas	0.080	14:43
05 Room Air	0.000	14:43
06 Std. Gas	0.080	14:44
07 Room Air	0.000	14:44

Lot No = 19817080A4  
Cyl No = 21  
Exp Date = 09/05/2019  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080 AC



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004946  
Location = TOXL      8164.14.00 09/16  
04/12/2018      14:44

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:45
02 Std. Gas	0.080	14:45
03 Room Air	0.000	14:46
04 Std. Gas	0.081	14:46
05 Room Air	0.000	14:47
06 Std. Gas	0.080	14:47
07 Room Air	0.000	14:47

Lot No = 19817080A4  
Cyl No = 21  
Exp Date = 09/05/2019  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

CALIBRATION CHECK  
0.080AC

Form 106-I8000