

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-006506

Location: TOXL

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 160 psi Regulator: 175 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check
 - a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: MP3063 Lot #: 16200 Exp. Date: 17 Aug 18
 - b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.300 AC
Sim. Ser #: MP3065 Lot #: 17350 Exp. Date: 11 Oct 19

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR5144 Lot #: 1C55 Exp. Date: 15 May 19
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 20016080A2 Cyl #: 48 Exp. Date: 5 Sep 18

Test 1 <u>0.082</u>	Test 4 <u>0.082</u>	Test 7 <u>0.082</u>
Test 2 <u>0.082</u>	Test 5 <u>0.082</u>	Test 8 <u>0.082</u>
Test 3 <u>0.081</u>	Test 6 <u>0.082</u>	Test 9 <u>0.082</u>
Average <u>0.082</u>		

C. Remarks/Maintenance: Annual inspection. Checked modern connection before and after performing annual inspection. Fax connection functioned properly both times.

Instrument is acceptable to be used in the field.

Robert Huggins
Breath Analyst Signature

17 Apr 18
Date

Charles E. Ed
Reviewed by

18 APR 18
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006506
Location = TOXL 8164.14.00 09/16
04/17/2018 14:31

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 20016080A2
Standard Cyl #? 48
Standard Expiration? 09/05/2018
Oper No? 888888

Flow Cal. Date: 08/20/2015
Slope 657
Intercept -627881

IR Calibration Date: 08/20/2015

	3um	9um
0th Coef(*100):	-22470	-18919
1st Coef(*100):	266505	136109
2nd Coef(*100):	3702	1644
H2O adj(mg/l*10k):	709	510

***** Printer Test End *****



Operator Signature
ROBERTA GRIEGER-NIMMO

Remarks: Print Test

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006506
Location = TOXL 8164.14.00 09/16
04/17/2018 14:55

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:56
02 Std. Sol.	0.021	14:57
03 Room Air	0.000	14:57
04 Std. Sol.	0.021	14:58
05 Room Air	0.000	14:58
06 Std. Sol.	0.021	14:59
07 Room Air	0.000	15:00

08 Sim Temp = 34.0°C

Simul Ser No = MP3063
Std Sol No = 16200
County = 08 Oper No. = 888888



Operator Signature
ROBERTA GRIEGER-NIMMO

Remarks: low AC
0.020 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006506
Location = TOXL 8164.14.00 09/16
04/17/2018 15:06

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:07
02 Std. Sol.	0.297	15:08
03 Room Air	0.000	15:09
04 Std. Sol.	0.298	15:10
05 Room Air	0.000	15:10
06 Std. Sol.	0.298	15:11
07 Room Air	0.000	15:12


08 Sim Temp = 34.0°C

Simul Ser No = MP3065

Std Sol No = 17350

County = 08

Oper No. = 888888



Operator Signature
ROBERTA GRIEGER-NIMMO

Remarks: High AC
0.300 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006506
Location = TOXL 8164.14.00 09/16
04/17/2018 14:33

Test	AC	Time
01 Room Air	0.000	14:34
02 *Subject Test	INT*	14:34
03 Room Air	0.000	14:35

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 20016080A2
Cyl No = 48
Expiration Date = 09/05/2018
County = 08 Oper No. = 888888



Operator Signature
ROBERTA GRIEGER-NIMMO

Remarks: *Interferent Check*
0.10 AC + 0.05% Acetone

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006506
Location = TOXL 8164.14.00 09/16
04/17/2018 14:38

Test	AC	Time
01 Diagnostic	OK	14:39
02 Room Air	0.000	14:39
03 *Subject Test	RFI*	14:40
04 Room Air	0.000	14:40

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 20016080A2
Cyl No = 48
Expiration Date = 09/05/2018
County = 08 Oper No. = 888888

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.


Operator Signature
ROBERTA GRIEGER-NIMMO

Remarks: RFI Check

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006506
Location = TOXL 8164.14.00 09/16
04/17/2018 14:41

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:42
02 Std. Gas	0.082	14:42
03 Room Air	0.000	14:43
04 Std. Gas	0.082	14:43
05 Room Air	0.000	14:43
06 Std. Gas	0.081	14:44
07 Room Air	0.000	14:44


Lot No = 20016080A2

Cyl No = 48

Exp Date = 09/05/2018

County = 08

Oper No. = 888888



Operator Signature
ROBERTA GRIEGER-NIMMO

Remarks: Calibration Check
0.080 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006506
Location = TOXL 8164.14.00 09/16
04/17/2018 14:45

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:45
02 Std. Gas	0.082	14:46
03 Room Air	0.000	14:46
04 Std. Gas	0.082	14:47
05 Room Air	0.000	14:47
06 Std. Gas	0.082	14:47
07 Room Air	0.000	14:48

Lot No = 20016080A2
Cyl No = 48
Exp Date = 09/05/2018
County = 08 Oper No. = 888888


Operator Signature
ROBERTA GRIEGER-NIMMO

Remarks: Calibration Check
0.080 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006506
Location = TOXL 8164.14.00 09/16
04/17/2018 14:49

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:49
02 Std. Gas	0.082	14:50
03 Room Air	0.000	14:50
04 Std. Gas	0.082	14:50
05 Room Air	0.000	14:51
06 Std. Gas	0.082	14:51
07 Room Air	0.000	14:52

Lot No = 20016080A2
Cyl No = 48
Exp Date = 09/05/2018
County = 08 Oper No. = 888888


Operator Signature
ROBERTA GRIEGER-NIMMO

Remarks: Calibration Check
0.080AC