



ETHANOL GAS STANDARD CYLINDER REPORT
 OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION
 Toxicology Section/Breath Alcohol Program
 SFN 59282 (10-2013)

Chemical Test Operator Name (Print) chris estes		
Location dcso ellendale		Intoxilyzer® Serial Number 80-005959 <i>3420</i>
Gas Lot Number 34917080a3	Gas Cylinder Number 013	Gas Expiration Date 02/05/2020

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 A. 0.081 AC
 B. 0.080 AC
 C. 0.081 AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>Chris Estes</i>	Date 5-01-18
Reviewed By (Crime Laboratory Use Only) <i>Roberta Suegar Thomas / RSN</i>	Date 5-08-18 / 6-13-18

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005953
Location = DCSO 8164.14.00 09/16
05/01/2018 14:35

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:36
02 Std. Gas	0.081	14:36
03 Room Air	0.000	14:37
04 Std. Gas	0.081	14:37
05 Room Air	0.000	14:38
06 Std. Gas	0.081	14:38
07 Room Air	0.000	14:38

Lot No = 349170^{80A3}~~00A3~~ *RAN* 13 Jun 18
Cyl No = 013
Exp Date = 02/05/2020
County = 11 Oper No. = 020504



Operator Signature
CHRIS ESTES

Remarks: