

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-005956

Location: OLIV

OLIVER COUNTY  
S.O.

A. General Setup and Checks:

1.  Diagnostic Tests Pass and Instrument in Ready Mode
2.  Breath tube heated
3.  Date, time and location code (Level 2,E). Re-set if necessary.
4.  Print test (Level 1,P). Sign and attach test record.
5.  Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
  - a. Display: 955 psi Regulator: 990 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check
  - a.  Low AC (Level 1,C): Known Value  $\leq 0.03$  AC: 0.020 AC  
Sim. Ser #: DR3377 Lot #: 201610A Exp. Date: 10-4-18
  - b.  High AC (Level 1,C): Known Value  $\geq 0.25$  AC: 0.300 AC  
Sim. Ser #: DR5191 Lot #: 17350 Exp. Date: 10-11-19

Configure dry gas standard for the following tests (Level 1,S).

2.  Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone  
Sim. Ser #: DR3451 Lot #: ICS#5 Exp. Date: 5-15-19  
 Display reads "Interferent Detect"
3.  RFI Check (CMS Mode)  Display reads "RFI Detect"
4.  Dry Calibration Check (Level 1,C): Known Value 0.080 AC  
Gas Cyl Lot #: 34917080A3 Cyl #: 34 Exp. Date: 2/5/20

Test 1	<u>0.077</u>	Test 4	<u>0.078</u>	Test 7	<u>0.078</u>
Test 2	<u>0.077</u>	Test 5	<u>0.078</u>	Test 8	<u>0.079</u>
Test 3	<u>0.078</u>	Test 6	<u>0.078</u>	Test 9	<u>0.078</u>
Average	<u>0.078</u>				

C. Remarks/Maintenance: ANNUAL INSPECTION, REPLACED  
GAS REGULATOR

Instrument is acceptable to be used in the field.

Charles E. Edr  
Breath Analyst Signature

5-4-2018  
Date

Roberta Muecke-Nemmi  
Reviewed by

5-9-2018  
Date



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005956  
Location = OLIV      8164.14.00 09/16  
05/04/2018      11:40

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:41
02 Std. Sol.	0.017	11:41
03 Room Air	0.000	11:42
04 Std. Sol.	0.018	11:43
05 Room Air	0.000	11:43
06 Std. Sol.	0.018	11:44
07 Room Air	0.000	11:44

08 Sim Temp = 34.0°C

Simul Ser No = DR3377  
Std Sol No = 201610A  
County = 33      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:      *Low AC*  
                    *0.020 AC*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005956  
Location = OLIV      8164.14.00 09/16  
05/04/2018      11:48

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:49
02 Std. Sol.	0.300	11:50
03 Room Air	0.000	11:50
04 Std. Sol.	0.301	11:51
05 Room Air	0.000	11:52
06 Std. Sol.	0.302	11:52
07 Room Air	0.000	11:53


08 Sim Temp = 34.0°C

Simul Ser No = DR5191

Std Sol No = 17350

County = 33

Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

HIGH AC

0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005956  
Location = OLIV      8164.14.00 09/16  
05/04/2018      11:37

Test	AC	Time
01 Room Air	0.000	11:37
02 *Subject Test	INT*	11:38
03 Room Air	0.000	11:38

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = INTERFERENT  
Dr. Lic. = ND/DIS821456  
Lot No = 34917080A3  
Cyl No = 34  
Expiration Date = 02/05/2020  
County = 33      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: *Interferent Check*  
*0.10AC + 0.05% ACETONE*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005956  
Location = OLIV                    8164.14.00 09/16  
05/04/2018                            11:34

Test	AC	Time
01 Diagnostic	OK	11:35
02 Room Air	0.000	11:35
03 *Subject Test	RFI*	11:35
04 Room Air	0.000	11:36

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = RFI CHECK  
Dr. Lic. = ND/DIS821456  
Lot No = 34917080A3  
Cyl No = 34  
Expiration Date = 02/05/2020  
County = 33                            Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks:            *RFI CHECK*


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005956  
Location = OLIV      8164.14.00 09/16  
05/04/2018      11:55

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:56
02 Std. Gas	0.077	11:56
03 Room Air	0.000	11:56
04 Std. Gas	0.077	11:57
05 Room Air	0.000	11:57
06 Std. Gas	0.078	11:58
07 Room Air	0.000	11:58

Lot No = 34917080A3  
Cyl No = 34  
Exp Date = 02/05/2020  
County = 33      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080 AC


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005956  
Location = OLIV      8164.14.00 09/16  
05/04/2018      11:59

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:59
02 Std. Gas	0.078	11:59
03 Room Air	0.000	12:00
04 Std. Gas	0.078	12:00
05 Room Air	0.000	12:01
06 Std. Gas	0.078	12:01
07 Room Air	0.000	12:02

Lot No = 34917080A3  
Cyl No = 34  
Exp Date = 02/05/2020  
County = 33      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080 AC



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005956  
Location = OLIV      8164.14.00 09/16  
05/04/2018      12:02

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:03
02 Std. Gas	0.078	12:03
03 Room Air	0.000	12:04
04 Std. Gas	0.079	12:04
05 Room Air	0.000	12:04
06 Std. Gas	0.078	12:05
07 Room Air	0.000	12:05

Lot No = 34917080A3  
Cyl No = 34  
Exp Date = 02/05/2020  
County = 33      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080 AC