

CHAIN OF CUSTODY
BREATH ALCOHOL INSTRUMENT

Instrument: INTOXILYZER 8000 Serial Number: 80-006493

Agency being assigned instrument: O.A.G. - CRIME LAB

Date 5-11-18

Transferred from:

Name (print) Danielle Bjorklie

Signature [Signature] 4482

Agency Nelson Co SO

Transferred to:

Name (print) CHARLES E. EDER

Signature [Signature]

Agency O.A.G. - CRIME LAB
(ANNUAL INSPECTION)

Reason for transfer/remarks:

- INSTRUMENT WAS RUNNING ABOVE $\pm 5\%$ FOR HIGH AC KNOWN VALUE
- INSTRUMENT WILL RETURN TO CRIME LAB FOR CALIBRATION ADJUSTMENT

Or Shipped Via:

USPS UPS FedEx Other

Date Sending Instrument: / / Air Bill No.