

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80004938

Location: LANG

CAVALIER COUNTY

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode S.O.
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 995 psi Regulator: 1000 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: DR3377 Lot #: 201610A Exp. Date: 10.4.18
- b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.300 AC
Sim. Ser #: DR5191 Lot #: 17350 Exp. Date: 10.11.19

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR3451 Lot #: ICS #5 Exp. Date: 5.15.19
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 20016080A2 Cyl #: 46 Exp. Date: 9.5.18

Test 1	<u>0.081</u>	Test 4	<u>0.082</u>	Test 7	<u>0.082</u>
Test 2	<u>0.081</u>	Test 5	<u>0.083</u>	Test 8	<u>0.083</u>
Test 3	<u>0.082</u>	Test 6	<u>0.083</u>	Test 9	<u>0.082</u>
Average	<u>0.082</u>				

C. Remarks/Maintenance: ANNUAL INSPECTION

Instrument is acceptable to be used in the field.

Charles E. Egan
Breath Analyst Signature

5.10.2018
Date

Roberto Henriquez-Munoz
Reviewed by

5.14.2018
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = LANG 8164.14.00 09/16
05/10/2018 10:47

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:48
02 Std. Sol.	0.019	10:49
03 Room Air	0.000	10:49
04 Std. Sol.	0.020	10:50
05 Room Air	0.000	10:50
06 Std. Sol.	0.020	10:51
07 Room Air	0.000	10:52

08 Sim Temp = 34.0°C

Simul Ser No = DR3377

Std Sol No = 201610A

County = 10

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

Low AC
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = LANG 8164.14.00 09/16
05/10/2018 10:53

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:54
02 Std. Sol.	0.299	10:55
03 Room Air	0.000	10:55
04 Std. Sol.	0.303	10:56
05 Room Air	0.000	10:57
06 Std. Sol.	0.305	10:57
07 Room Air	0.000	10:58

08 Sim Temp = 34.0°C

Simul Ser No = DR5191
Std Sol No = 17350
County = 10 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

HIGH AC
0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = LANG 8164.14.00 09/16
05/10/2018 10:35

Test	AC	Time
01 Room Air	0.000	10:36
02 *Subject Test	INT*	10:37
03 Room Air	0.000	10:38

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 03/01/1982
Sub Sex = Male Weight = 120
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821449
Lot No = 20016080A2
Cyl No = 46
Expiration Date = 09/05/2018
County = 10 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent Check

0.10AC + 0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = LANG 8164.14.00 09/16
05/10/2018 10:32

Test	AC	Time
01 Diagnostic	OK	10:33
02 Room Air	0.000	10:34
03 *Subject Test	RFI*	10:34
04 Room Air	0.000	10:34

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 03/01/1982
Sub Sex = Male Weight = 120
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821449
Lot No = 20016080A2
Cyl No = 46
Expiration Date = 09/05/2018
County = 10 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = LANG 8164.14.00 09/16
05/10/2018 10:39

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:39
02 Std. Gas	0.081	10:39
03 Room Air	0.000	10:40
04 Std. Gas	0.081	10:40
05 Room Air	0.000	10:41
06 Std. Gas	0.082	10:41
07 Room Air	0.000	10:42

Lot No = 20016080A2
Cyl No = 46
Exp Date = 09/05/2018
County = 10 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = LANG 8164.14.00 09/16
05/10/2018 10:42

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:43
02 Std. Gas	0.082	10:44
03 Room Air	0.000	10:44
04 Std. Gas	0.083	10:44
05 Room Air	0.000	10:45
06 Std. Gas	0.083	10:45
07 Room Air	0.000	10:46

Lot No = 20016080A2
Cyl No = 46
Exp Date = 09/05/2018
County = 10 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = LANG 8164.14.00 09/16
05/10/2018 11:00

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:00
02 Std. Gas	0.082	11:00
03 Room Air	0.000	11:01
04 Std. Gas	0.083	11:01
05 Room Air	0.000	11:02
06 Std. Gas	0.082	11:02
07 Room Air	0.000	11:03

Lot No = 20016080A2
Cyl No = 46
Exp Date = 09/05/2018
County = 10 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000