

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-004940

Location: FTOT

BIA - FORT TOTTEN

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 1012 psi Regulator: 1000 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a. Low AC (Level 1,C): Known Value \leq 0.03 AC: 0.020 AC
Sim. Ser #: DR3377 Lot #: 2016104 Exp. Date: 10.9.18
- b. High AC (Level 1,C): Known Value \geq 0.25 AC: 0.300 AC
Sim. Ser #: DR5191 Lot #: 17350 Exp. Date: 10.11.19

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR3451 Lot #: ICS # 5 Exp. Date: 5.15.19
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 34917080A3 Cyl #: 2 Exp. Date: 2/5/20

Test 1	<u>0.080</u>	Test 4	<u>0.081</u>	Test 7	<u>0.081</u>
Test 2	<u>0.080</u>	Test 5	<u>0.080</u>	Test 8	<u>0.081</u>
Test 3	<u>0.080</u>	Test 6	<u>0.080</u>	Test 9	<u>0.081</u>
Average	<u>0.080</u>				

C. Remarks/Maintenance: ANNUAL INSPECTION, REPLACED SIMULATOR
RETURN O-RING

Instrument is acceptable to be used in the field.

Charles E. Ed
Breath Analyst Signature

5-9-2018
Date

Rodolfo Murgueta-Nemmo
Reviewed by

5-14-2018
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004940
Location = FTOT 8164.14.00 09/16
05/09/2018 17:24

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxy1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 34917080A3
Standard Cyl #? 2
Standard Expiration? 02/05/2020
Oper No? 666666

Flow Cal. Date: 01/18/2010
Slope 720
Intercept -746294

IR Calibration Date: 06/01/2015
 3um 9um

0th Coef(*100): -14429 -15122
1st Coef(*100): 263440 134864
2nd Coef(*100): 2296 1357
H2O adj(mg/l*10k): 371 332

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004940
Location = FTOT 8164.14.00 09/16
05/09/2018 17:49

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:50
02 Std. Sol.	0.019	17:51
03 Room Air	0.000	17:51
04 Std. Sol.	0.020	17:52
05 Room Air	0.000	17:53
06 Std. Sol.	0.019	17:53
07 Room Air	0.000	17:54

08 Sim Temp = 34.0°C

Simul Ser No = DR3377
Std Sol No = 201610A
County = 62 Oper No. = 666666

CEE
X Charles E. Eder

Operator Signature
CHARLES EDER

Remarks: *Low AC*
0.020 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004940
Location = FTOT 8164.14.00 09/16
05/09/2018 17:44

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:45
02 Std. Sol.	0.300	17:45
03 Room Air	0.000	17:46
04 Std. Sol.	0.299	17:47
05 Room Air	0.000	17:47
06 Std. Sol.	0.299	17:48
07 Room Air	0.000	17:48


08 Sim Temp = 34.0°C

Simul Ser No = DR5191

Std Sol No = 17350

County = 62

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

HIGH AC
0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004940
Location = FTOT 8164.14.00 09/16
05/09/2018 17:28

Test	AC	Time
01 Room Air	0.000	17:29
02 *Subject Test	INT*	17:29
03 Room Air	0.000	17:30

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 03/01/1982
Sub Sex = Male Weight = 120
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821449
Lot No = 34917080A3
Cyl No = 2
Expiration Date = 02/05/2020
County = 62 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent Check
0.10AC + 0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004940
Location = FTOT 8164.14.00 09/16
05/09/2018 17:25

Test	AC	Time
01 Diagnostic	OK	17:26
02 Room Air	0.000	17:27
03 *Subject Test	RFI*	17:27
04 Room Air	0.000	17:27

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 03/01/1982
Sub Sex = Male Weight = 120
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821449
Lot No = 34917080A3
Cyl No = 2
Expiration Date = 02/05/2020
County = 62 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004940
Location = FTOT 8164.14.00 09/16
05/09/2018 17:31

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:32
02 Std. Gas	0.080	17:32
03 Room Air	0.000	17:32
04 Std. Gas	0.080	17:33
05 Room Air	0.000	17:33
06 Std. Gas	0.080	17:34
07 Room Air	0.000	17:34

Lot No = 34917080A3

Cyl No = 2

Exp Date = 02/05/2020

County = 62

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION CHECK
0.080AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004940
Location = FTOT 8164.14.00 09/16
05/09/2018 17:34

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:35
02 Std. Gas	0.081	17:35
03 Room Air	0.000	17:36
04 Std. Gas	0.080	17:36
05 Room Air	0.000	17:37
06 Std. Gas	0.080	17:37
07 Room Air	0.000	17:37

Lot No = 34917080A3
Cyl No = 2
Exp Date = 02/05/2020
County = 62 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *CALIBRATION CHECK*
0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004940
Location = FTOT 8164.14.00 09/16
05/09/2018 17:58

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:58
02 Std. Gas	0.081	17:59
03 Room Air	0.000	17:59
04 Std. Gas	0.081	18:00
05 Room Air	0.000	18:00
06 Std. Gas	0.081	18:01
07 Room Air	0.000	18:01

Lot No = 34917080A3
Cyl No = 2
Exp Date = 02/05/2020
County = 62 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION CHECK
0.080 AC

Form 106-I8000