

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-005361

Location: M050

MOUNTAIN COUNTY  
JAIL  
CORRECTIONAL CENTER

A. General Setup and Checks:

1.  Diagnostic Tests Pass and Instrument in Ready Mode
2.  Breath tube heated
3.  Date, time and location code (Level 2,E). Re-set if necessary.
4.  Print test (Level 1,P). Sign and attach test record.
5.  Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
  - a. Display: 990 psi Regulator: 1060 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a.  Low AC (Level 1,C): Known Value  $\leq 0.03$  AC: 0.020 AC  
Sim. Ser #: DR3377 Lot #: 201610A Exp. Date: 10-4-18
- b.  High AC (Level 1,C): Known Value  $\geq 0.25$  AC: 0.300 AC  
Sim. Ser #: DR3191 Lot #: 17350 Exp. Date: 10-11-19

Configure dry gas standard for the following tests (Level 1,S).

2.  Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone  
Sim. Ser #: DR3451 Lot #: ICS #5 Exp. Date: 5.15.19

Display reads "Interferent Detect"

3.  RFI Check (CMS Mode)  Display reads "RFI Detect"

4.  Dry Calibration Check (Level 1,C): Known Value 0.080 AC  
Gas Cyl Lot #: 03417080A3 Cyl #: 2 Exp. Date: 3-5-19

Test 1	<u>0.081</u>	Test 4	<u>0.081</u>	Test 7	<u>0.081</u>
Test 2	<u>0.082</u>	Test 5	<u>0.081</u>	Test 8	<u>0.081</u>
Test 3	<u>0.082</u>	Test 6	<u>0.082</u>	Test 9	<u>0.081</u>
Average	<u>0.081</u>				

C. Remarks/Maintenance: ANNUAL INSPECTION, REPLACED  
SIMULATOR RETURN O-RING.

Instrument is acceptable to be used in the field.

Charles E. Ed  
Breath Analyst Signature

5.8.2018  
Date

Robert H. Huggins - Thomas  
Reviewed by

14 May 18  
Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005361  
Location = MOSO                    8164.14.00 09/16  
05/08/2018                            12:22

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

Current Instrument Setup


Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?          1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            03417080A3  
Standard Cyl #?            2  
Standard Expiration?       03/05/2019  
Oper No?                    666666

Flow Cal. Date:            08/19/2011  
Slope                        645  
Intercept                   -487053

IR Calibration Date:        01/13/2012  
                                  3um                            9um

-----  
0th Coef(\*100):            -5818                        -10599  
1st Coef(\*100):            251241                       123805  
2nd Coef(\*100):            7259                         2455  
H2O adj(mg/l\*10k):        715                          798

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005361  
Location = MOSO      8164.14.00 09/16  
05/08/2018      12:41

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:42
02 Std. Sol.	0.018	12:43
03 Room Air	0.000	12:43
04 Std. Sol.	0.018	12:44
05 Room Air	0.000	12:44
06 Std. Sol.	0.018	12:45
07 Room Air	0.000	12:46

08 Sim Temp = 34.0°C

Simul Ser No = DR3377

Std Sol No = 201610A

County = 31

Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

Low AC  
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005361  
Location = MOSO      8164.14.00 09/16  
05/08/2018      12:56

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:56
02 Std. Sol.	0.296	12:57
03 Room Air	0.000	12:58
04 Std. Sol.	0.294	12:58
05 Room Air	0.000	12:59
06 Std. Sol.	0.295	13:00
07 Room Air	0.000	13:00

08 Sim Temp = 34.0°C

Simul Ser No = DR5191  
Std Sol No = 17350  
County = 31      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: *HIGH AC*  
*cee ~~0.00~~ 0.300 AC*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005361  
Location = MOSO                8164.14.00 09/16  
05/08/2018                        12:31

Test	AC	Time
01 Room Air	0.000	12:32
02 *Subject Test	INT*	12:33
03 Room Air	0.000	12:33

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 03/01/1982  
Sub Sex = Male                    Weight = 120  
Test = DUI                        Cit = INTERFERENT  
Dr. Lic. = ND/DIS821449  
Lot No = 03417080A3  
Cyl No = 2  
Expiration Date = 03/05/2019  
County = 31                        Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: *Interferent Check*  
*0.10 AC + 0.05% ACETONE*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005361  
Location = MOSO                    8164.14.00 09/16  
05/08/2018                            12:28

Test	AC	Time
01 Diagnostic	OK	12:30
02 Room Air	0.000	12:30
03 *Subject Test	RFI*	12:30
04 Room Air	0.000	12:31

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 03/01/1982  
Sub Sex = Male                            Weight = 120  
Test = DUI                                Cit = RFI CHECK  
Dr. Lic. = ND/DIS821449  
Lot No = 03417080A3  
Cyl No = 2  
Expiration Date = 03/05/2019  
County = 31                                Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks:            *RFI CHECK*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005361  
Location = MOSO      8164.14.00 09/16  
05/08/2018      12:34

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:34
02 Std. Gas	0.081	12:35
03 Room Air	0.000	12:35
04 Std. Gas	0.082	12:36
05 Room Air	0.000	12:36
06 Std. Gas	0.082	12:36
07 Room Air	0.000	12:37


Lot No = 03417080A3

Cyl No = 2

Exp Date = 03/05/2019

County = 31

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

CALIBRATION CHECK

0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005361  
Location = MOSO      8164.14.00 09/16  
05/08/2018      12:37

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:38
02 Std. Gas	0.081	12:38
03 Room Air	0.000	12:39
04 Std. Gas	0.081	12:39
05 Room Air	0.000	12:40
06 Std. Gas	0.082	12:40
07 Room Air	0.000	12:41

Lot No = 03417080A3  
Cyl No = 2  
Exp Date = 03/05/2019  
County = 31      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

CALIBRATION CHECK  
0.080 AC

Form 106-I8000



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005361  
Location = MOSO      8164.14.00 09/16  
05/08/2018      13:01

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:01
02 Std. Gas	0.081	13:02
03 Room Air	0.000	13:02
04 Std. Gas	0.081	13:03
05 Room Air	0.000	13:03
06 Std. Gas	0.081	13:04
07 Room Air	0.000	13:04

Lot No = 03417080A3  
Cyl No = 2  
Exp Date = 03/05/2019  
County = 31      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080 AC