

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-006505

Location: MOHL

A. General Setup and Checks:

RENNVILLE COUNTY  
S.O.

1.  Diagnostic Tests Pass and Instrument in Ready Mode
2.  Breath tube heated
3.  Date, time and location code (Level 2,E). Re-set if necessary.
4.  Print test (Level 1,P). Sign and attach test record.
5.  Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
  - a. Display: 699 psi Regulator: 725 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check
  - a.  Low AC (Level 1,C): Known Value  $\leq 0.03$  AC: 0.020 AC  
Sim. Ser #: DR3377 Lot #: 201610A Exp. Date: 10-4-18
  - b.  High AC (Level 1,C): Known Value  $\geq 0.25$  AC: 0.300 AC  
Sim. Ser #: DR5191 Lot #: 17350 Exp. Date: 10-11-19

Configure dry gas standard for the following tests (Level 1,S).

2.  Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone  
Sim. Ser #: DR351EE Lot #: ICS#5 Exp. Date: 5-15-19  
 Display reads "Interferent Detect"
3.  RFI Check (CMS Mode)  Display reads "RFI Detect"
4.  Dry Calibration Check (Level 1,C): Known Value 0.080 AC  
Gas Cyl Lot #: 17316080A2 Cyl #: 31 Exp. Date: 8/5/18

Test 1 <u>0.081</u>	Test 4 <u>0.080</u>	Test 7 <u>0.080</u>
Test 2 <u>0.080</u>	Test 5 <u>0.080</u>	Test 8 <u>0.080</u>
Test 3 <u>0.080</u>	Test 6 <u>0.080</u>	Test 9 <u>0.080</u>
Average <u>0.080</u>		

C. Remarks/Maintenance: ANNUAL INSPECTION

Instrument is acceptable to be used in the field.

Charles E. Ede  
Breath Analyst Signature

5-9-2018  
Date

Roberta Huong-Turmon  
Reviewed by

14 May 18  
Date



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006505  
Location = MOHL      8164.14.00 09/16  
05/09/2018      09:37

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:38
02 Std. Sol.	0.017	09:38
03 Room Air	0.000	09:39
04 Std. Sol.	0.018	09:40
05 Room Air	0.000	09:40
06 Std. Sol.	0.017	09:41
07 Room Air	0.000	09:41

08 Sim Temp = 34.0°C

Simul Ser No = DR3377

Std Sol No = 201610A

County = 38

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

Low AC  
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006505  
Location = MOHL      8164.14.00 09/16  
05/09/2018      09:42

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:43
02 Std. Sol.	0.304	09:44
03 Room Air	0.000	09:44
04 Std. Sol.	0.308	09:45
05 Room Air	0.000	09:46
06 Std. Sol.	0.308	09:46
07 Room Air	0.000	09:47

08 Sim Temp = 34.0°C

Simul Ser No = DR5191

Std Sol No = 17350

County = 38

Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

HGA AC

0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-006505  
Location = MOHL                    8164.14.00 09/16  
05/09/2018                            09:14

Test	AC	Time
01 Room Air	0.000	09:15
02 *Subject Test	INT*	09:16
03 Room Air	0.000	09:17

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 03/01/1982  
Sub Sex = Male                            Weight = 120  
Test = DUI                                Cit = INTERFERENT  
Dr. Lic. = ND/DIS821449  
Lot No = 17316080A2  
Cyl No = 31  
Expiration Date = 08/05/2018  
County = 38                                Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*Interferent Check*  
*0.10 AC + 0.05% ACETONE*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer                      Alcohol Analyzer  
North Dakota Model 8000                      SN 80-006505  
Location = MOHL                              8164.14.00 09/16  
05/09/2018    09:11

Test	AC	Time
01 Diagnostic	OK	09:12
02 Room Air	0.000	09:12
03 *Subject Test	RFI*	09:13
04 Room Air	0.000	09:13

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 03/01/1982  
Sub Sex = Male                                      Weight = 120  
Test = DUI    Cit = RFI CHECK  
Dr. Lic. = ND/DIS821449  
Lot No = 17316080A2  
Cyl No = 31  
Expiration Date = 08/05/2018  
County = 38    Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks:

*RFI CHECK*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006505  
Location = MOHL      8164.14.00 09/16  
05/09/2018      09:17

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:17
02 Std. Gas	0.081	09:18
03 Room Air	0.000	09:18
04 Std. Gas	0.080	09:19
05 Room Air	0.000	09:19
06 Std. Gas	0.080	09:20
07 Room Air	0.000	09:20

Lot No = 17316080A2  
Cyl No = 31  
Exp Date = 08/05/2018  
County = 38      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080 AC

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006505  
Location = MOHL      8164.14.00 09/16  
05/09/2018      09:22

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:22
02 Std. Gas	0.080	09:23
03 Room Air	0.000	09:23
04 Std. Gas	0.080	09:23
05 Room Air	0.000	09:24
06 Std. Gas	0.080	09:24
07 Room Air	0.000	09:25

Lot No = 17316080A2  
Cyl No = 31  
Exp Date = 08/05/2018  
County = 38      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK

0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006505  
Location = MOHL      8164.14.00 09/16  
05/09/2018      09:47

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:48
02 Std. Gas	0.080	09:48
03 Room Air	0.000	09:49
04 Std. Gas	0.080	09:49
05 Room Air	0.000	09:50
06 Std. Gas	0.080	09:50
07 Room Air	0.000	09:51

Lot No = 17316080A2  
Cyl No = 31  
Exp Date = 08/05/2018  
County = 38      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080 AC

Form 106-I8000