

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-005942

Location: POWERS LAKE
P.D (PLPD)

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: ~~1000~~ 1014 psi Regulator: 1000 psi
CEE

B. Tests (Sign and attach test records):

- Configure simulator for the following tests (Level 1,S).
1. Wet Calibration Check
 - a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: DR3377 Lot #: 2016 104 Exp. Date: 10-9-18
 - b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.300 AC
Sim. Ser #: DR5191 Lot #: 17350 Exp. Date: 10-11-19

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR3451 Lot #: ICS #5 Exp. Date: 5-15-19
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 03417080A3 Cyl #: 5 Exp. Date: 3-5-19

Test 1	<u>0.079</u>	Test 4	<u>0.079</u>	Test 7	<u>0.078</u>
Test 2	<u>0.080</u>	Test 5	<u>0.080</u>	Test 8	<u>0.079</u>
Test 3	<u>0.079</u>	Test 6	<u>0.080</u>	Test 9	<u>0.079</u>
Average <u>0.079</u>					

C. Remarks/Maintenance: ANNUAL INSPECTION

Instrument is acceptable to be used in the field.

Charles E. Eder
Breath Analyst Signature

5-8-18
Date

Roberta Hugu-Thomas
Reviewed by

14 May 18
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005942
Location = PLPD 8164.14.00 09/16
05/08/2018 10:54

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 03417080A3
Standard Cyl #? 5
Standard Expiration? 03/05/2019
Oper No? 666666

Flow Cal. Date: 06/18/2013
Slope 715
Intercept -740740

IR Calibration Date: 05/20/2015
 3um 9um

0th Coef(*100): -32049 -32893
1st Coef(*100): 272377 139703
2nd Coef(*100): 2023 1232
H2O adj(mg/l*10k): 468 428

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005942
Location = PLPD 8164.14.00 09/16
05/08/2018 11:29

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:30
02 Std. Sol.	0.016	11:31
03 Room Air	0.000	11:31
04 Std. Sol.	0.016	11:32
05 Room Air	0.000	11:32
06 Std. Sol.	0.016	11:33
07 Room Air	0.000	11:34


08 Sim Temp = 34.0°C

Simul Ser No = DR3377

Std Sol No = 201610A

County = 07

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Low AC

0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005942
Location = PLPD 8164.14.00 09/16
05/08/2018 11:35

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:35
02 Std. Sol.	0.297	11:36
03 Room Air	0.000	11:37
04 Std. Sol.	0.299	11:37
05 Room Air	0.000	11:38
06 Std. Sol.	0.300	11:39
07 Room Air	0.000	11:39

08 Sim Temp = 34.0°C

Simul Ser No = DR5191

Std Sol No = 17350

County = 07

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

HIGH AC
0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005942
Location = PLPD 8164.14.00 09/16
05/08/2018 10:59

Test	AC	Time
01 Room Air	0.000	11:00
02 *Subject Test	INT*	11:01
03 Room Air	0.000	11:01

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 03/01/1982
Sub Sex = Male Weight = 120
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821449
Lot No = 03417080A3
Cyl No = 5
Expiration Date = 03/05/2019
County = 07 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent Check

0.10 AC + 0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005942
Location = PLPD 8164.14.00 09/16
05/08/2018 10:56

Test	AC	Time
01 Diagnostic	OK	10:57
02 Room Air	0.000	10:57
03 *Subject Test	RFI*	10:57
04 Room Air	0.000	10:58

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 03/01/1982
Sub Sex = Male Weight = 120
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821449
Lot No = 03417080A3
Cyl No = 5
Expiration Date = 03/05/2019
County = 07 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: *RFI CHECK*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005942
Location = PLPD 8164.14.00 09/16
05/08/2018 11:03

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:03
02 Std. Gas	0.079	11:03
03 Room Air	0.000	11:04
04 Std. Gas	0.080	11:04
05 Room Air	0.000	11:05
06 Std. Gas	0.079	11:05
07 Room Air	0.000	11:06

Lot No = 03417080A3
Cyl No = 5
Exp Date = 03/05/2019
County = 07 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *CALIBRATION CHECK*
0.080 AC


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005942
Location = PLPD 8164.14.00 09/16
05/08/2018 11:06

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:07
02 Std. Gas	0.079	11:07
03 Room Air	0.000	11:08
04 Std. Gas	0.080	11:08
05 Room Air	0.000	11:09
06 Std. Gas	0.080	11:09
07 Room Air	0.000	11:09

Lot No = 03417080A3
Cyl No = 5
Exp Date = 03/05/2019
County = 07 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005942
Location = PLPD 8164.14.00 09/16
05/08/2018 11:40

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:41
02 Std. Gas	0.078	11:41
03 Room Air	0.000	11:42
04 Std. Gas	0.079	11:42
05 Room Air	0.000	11:43
06 Std. Gas	0.079	11:43
07 Room Air	0.000	11:43

Lot No = 03417080A3
Cyl No = 5
Exp Date = 03/05/2019
County = 07 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *CALIBRATION CHECK*
0.080AC