

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-005946

Location: WILLIAMS COUNTY L.E.C.

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 575 psi Regulator: 550 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: DR3377 Lot #: 201610A Exp. Date: 10-4-18
- b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.300 AC
Sim. Ser #: DR5191 Lot #: 17350 Exp. Date: 10-11-19

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR351 Lot #: JCS #5 Exp. Date: 5-15-19
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 0341708043 Cyl #: 34 Exp. Date: 3-5-19

Test 1	<u>0.081</u>	Test 4	<u>0.079</u>	Test 7	<u>0.080</u>
Test 2	<u>0.079</u>	Test 5	<u>0.079</u>	Test 8	<u>0.078</u>
Test 3	<u>0.080</u>	Test 6	<u>0.079</u>	Test 9	<u>0.079</u>
Average	<u>0.079</u>				

C. Remarks/Maintenance: ANNUAL INSPECTION, FIXED TORN COVER ON BREATH HOSE (Electrical tape), REPLACED SIMULATOR RETURN O-RING

Instrument is acceptable to be used in the field.

Charles E. [Signature]
Breath Analyst Signature

5-7-2018
Date

Roberta Mungu-Nimmo
Reviewed by

14 May 18
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005946
Location = WILL 8164.14.00 09/16
05/07/2018 15:51

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-|=|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-|=|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 03417080A3
Standard Cyl #? 34
Standard Expiration? 03/05/2019
Oper No? 666666

Flow Cal. Date: 08/10/2015
Slope 673
Intercept -638348

IR Calibration Date: 08/10/2015
3um 9um

0th Coef(*100): -22507 -13665
1st Coef(*100): 273651 130615
2nd Coef(*100): 5249 2073
H2O adj(mg/l*10k): 617 504

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks: PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005946
Location = WILL 8164.14.00 09/16
05/07/2018 16:16

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:16
02 Std. Sol.	0.018	16:17
03 Room Air	0.000	16:18
04 Std. Sol.	0.017	16:18
05 Room Air	0.000	16:19
06 Std. Sol.	0.017	16:20
07 Room Air	0.000	16:20

08 Sim Temp = 34.0°C

Simul Ser No = DR3377

Std Sol No = 201610A

County = 53

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Low AC

0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005946
Location = WILL 8164.14.00 09/16
05/07/2018 16:21

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:21
02 Std. Sol.	0.299	16:22
03 Room Air	0.000	16:23
04 Std. Sol.	0.304	16:23
05 Room Air	0.000	16:24
06 Std. Sol.	0.303	16:25
07 Room Air	0.000	16:25

08 Sim Temp = 34.0°C

Simul Ser No = DR5191
Std Sol No = 17350
County = 53 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: HIGH AC
0.300 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005946
Location = WILL 8164.14.00 09/16
05/07/2018 16:01

Test	AC	Time
01 Room Air	0.000	16:02
02 *Subject Test	INT*	16:02
03 Room Air	0.000	16:03

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 03/01/1982
Sub Sex = Male Weight = 120
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821449
Lot No = 03417080A3
Cyl No = 34
Expiration Date = 03/05/2019
County = 53 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent Check

0.10 AC + 0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005946
Location = WILL 8164.14.00 09/16
05/07/2018 15:58

Test	AC	Time
01 Diagnostic	OK	15:59
02 Room Air	0.000	16:00
03 *Subject Test	RFI*	16:00
04 Room Air	0.000	16:01

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 03/01/1982
Sub Sex = Male Weight = 120
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821449
Lot No = 03417080A3
Cyl No = 34
Expiration Date = 03/05/2019
County = 53 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005946
Location = WILL 8164.14.00 09/16
05/07/2018 16:04

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:04
02 Std. Gas	0.081	16:05
03 Room Air	0.000	16:05
04 Std. Gas	0.079	16:06
05 Room Air	0.000	16:06
06 Std. Gas	0.080	16:07
07 Room Air	0.000	16:07

Lot No = 03417080A3

Cyl No = 34

Exp Date = 03/05/2019

County = 53

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005946
Location = WILL 8164.14.00 09/16
05/07/2018 16:07

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:08
02 Std. Gas	0.079	16:08
03 Room Air	0.000	16:09
04 Std. Gas	0.079	16:09
05 Room Air	0.000	16:10
06 Std. Gas	0.079	16:10
07 Room Air	0.000	16:11

Lot No = 03417080A3

Cyl No = 34

Exp Date = 03/05/2019

County = 53

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION CHECK

0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005946
Location = WILL 8164.14.00 09/16
05/07/2018 16:26

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:26
02 Std. Gas	0.080	16:27
03 Room Air	0.000	16:27
04 Std. Gas	0.078	16:28
05 Room Air	0.000	16:28
06 Std. Gas	0.079	16:29
07 Room Air	0.000	16:29

Lot No = 03417080A3
Cyl No = 34
Exp Date = 03/05/2019
County = 53 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC