

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-005954

Location: DEVL

LAKE REGION L.E.C.

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 1012 psi Regulator: 1000 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check
 - a. Low AC (Level 1,C): Known Value \leq 0.03 AC: 0.020 AC
Sim. Ser #: DR3377 Lot #: 201610A Exp. Date: 10-4-18
 - b. High AC (Level 1,C): Known Value \geq 0.25 AC: 0.380 AC
Sim. Ser #: DR5191 Lot #: 17350 Exp. Date: 10-11-19

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR3451 Lot #: ICS#5 Exp. Date: 5-15-19
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 34917080A3 Cyl #: 3 Exp. Date: 2-5-20

Test 1	<u>0.078</u>	Test 4	<u>0.080</u>	Test 7	<u>0.079</u>
Test 2	<u>0.079</u>	Test 5	<u>0.079</u>	Test 8	<u>0.079</u>
Test 3	<u>0.079</u>	Test 6	<u>0.079</u>	Test 9	<u>0.079</u>
Average	<u>0.079</u>				

C. Remarks/Maintenance: ANNUAL INSPECTION

Instrument is acceptable to be used in the field.

Charles E. Eda
Breath Analyst Signature

5-9-2018
Date

Robert J. Huggins-Tumino
Reviewed by

14 May 18
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005954
Location = DEVL 8164.14.00 09/16
05/09/2018 16:45

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:45
02 Std. Sol.	0.019	16:46
03 Room Air	0.000	16:47
04 Std. Sol.	0.019	16:47
05 Room Air	0.000	16:48
06 Std. Sol.	0.019	16:49
07 Room Air	0.000	16:49

08 Sim Temp = 34.0°C

Simul Ser No = DR3377
Std Sol No = 201610A
County = 36 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Low AC*
0.020 AC
Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005954
Location = DEVL 8164.14.00 09/16
05/09/2018 16:39

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:40
02 Std. Sol.	0.300	16:41
03 Room Air	0.000	16:41
04 Std. Sol.	0.299	16:42
05 Room Air	0.000	16:43
06 Std. Sol.	0.299	16:43
07 Room Air	0.000	16:44

08 Sim Temp = 34.0°C

Simul Ser No = DR5191

Std Sol No = 17350

County = 36

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

HIGH AC
0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005954
Location = DEVL 8164.14.00 09/16
05/09/2018 16:16

Test	AC	Time
01 Room Air	0.000	16:17
02 *Subject Test	INT*	16:17
03 Room Air	0.000	16:18

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 03/01/1982
Sub Sex = Male Weight = 120
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821449
Lot No = 34917080A3
Cyl No = 3
Expiration Date = 02/05/2020
County = 36 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent Check
0.10 AC + 0.05% ACETONE


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005954
Location = DEVL 8164.14.00 09/16
05/09/2018 16:18

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:19
02 Std. Gas	0.078	16:19
03 Room Air	0.000	16:20
04 Std. Gas	0.079	16:20
05 Room Air	0.000	16:21
06 Std. Gas	0.079	16:21
07 Room Air	0.000	16:22

Lot No = 34917080A3
Cyl No = 3
Exp Date = 02/05/2020
County = 36 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005954
Location = DEVL 8164.14.00 09/16
05/09/2018 16:22

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:23
02 Std. Gas	0.080	16:23
03 Room Air	0.000	16:24
04 Std. Gas	0.079	16:24
05 Room Air	0.000	16:25
06 Std. Gas	0.079	16:25
07 Room Air	0.000	16:26

Lot No = 34917080A3
Cyl No = 3
Exp Date = 02/05/2020
County = 36 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005954
Location = DEVL 8164.14.00 09/16
05/09/2018 16:50

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:50
02 Std. Gas	0.079	16:51
03 Room Air	0.000	16:51
04 Std. Gas	0.079	16:52
05 Room Air	0.000	16:52
06 Std. Gas	0.079	16:53
07 Room Air	0.000	16:53

Lot No = 34917080A3
Cyl No = 3
Exp Date = 02/05/2020
County = 36 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION CHECK
0.080 AC

Form 106-I8000