

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-005957

Location: CROS
DIVIDE COUNTY
S.D.

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 1004 psi Regulator: 1000 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: DR3377 Lot #: 201610A Exp. Date: 10-4-18
- b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.300 AC
Sim. Ser #: DR5191 Lot #: 17350 Exp. Date: 10-11-19

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR3451 Lot #: ICS #5 Exp. Date: 5-15-19
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 8 17316080A2 Cyl #: 17 Exp. Date: 8/5/18

Test 1	<u>0.079</u>	Test 4	<u>0.079</u>	Test 7	<u>0.079</u>
Test 2	<u>0.080</u>	Test 5	<u>0.079</u>	Test 8	<u>0.079</u>
Test 3	<u>0.079</u>	Test 6	<u>0.079</u>	Test 9	<u>0.079</u>
Average <u>0.079</u>					

C. Remarks/Maintenance: ANNUAL INSPECTION, REPLACED
EXTERNAL LASER PRINTER (ERROR CODE ON PRINTER)

Instrument is acceptable to be used in the field.

Charles E. Edr
Breath Analyst Signature

5-8-2018
Date

Roberta Huequi-Nimmo
Reviewed by

14 May 18
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005957
Location = CROS 8164.14.00 09/16
05/08/2018 14:50

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-=_|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%&^*()_+?

abcdefghijklmnopqrstuvwxy1234567890-=_|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%&^*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 17316080A2
Standard Cyl #? 17
Standard Expiration? 08/05/2018
Oper No? 666666

Flow Cal. Date: 06/21/2013
Slope 647
Intercept -613317

IR Calibration Date: 05/11/2015
 3um 9um

0th Coef(*100): -16666 -18045
1st Coef(*100): 278478 132949
2nd Coef(*100): 3303 1397
H2O adj(mg/l*10k): 379 351

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005957
Location = CROS 8164.14.00 09/16
05/08/2018 15:12

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:13
02 Std. Sol.	0.017	15:14
03 Room Air	0.000	15:14
04 Std. Sol.	0.017	15:15
05 Room Air	0.000	15:16
06 Std. Sol.	0.018	15:16
07 Room Air	0.000	15:17


08 Sim Temp = 34.0°C

Simul Ser No = DR3377

Std Sol No = 201610A

County = 12

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Low AC
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005957
Location = CROS 8164.14.00 09/16
05/08/2018 15:18

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:19
02 Std. Sol.	0.297	15:20
03 Room Air	0.000	15:20
04 Std. Sol.	0.299	15:21
05 Room Air	0.000	15:22
06 Std. Sol.	0.300	15:22
07 Room Air	0.000	15:23

08 Sim Temp = 34.0°C

Simul Ser No = DR5191
Std Sol No = 17350
County = 12 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: HIGH AC
 0.300 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005957
Location = CROS 8164.14.00 09/16
05/08/2018 15:00

Test	AC	Time
01 Room Air	0.000	15:01
02 *Subject Test	INT*	15:01
03 Room Air	0.000	15:02

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 03/01/1982
Sub Sex = Male Weight = 120
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821449
Lot No = 17316080A2
Cyl No = 17
Expiration Date = 08/05/2018
County = 12 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent Check
0.10 AC + 0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005957
Location = CROS 8164.14.00 09/16
05/08/2018 14:58

Test	AC	Time
01 Diagnostic	OK	14:59
02 Room Air	0.000	14:59
03 *Subject Test	RFI*	14:59
04 Room Air	0.000	15:00

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 03/01/1982
Sub Sex = Male Weight = 120
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821449
Lot No = 17316080A2
Cyl No = 17
Expiration Date = 08/05/2018
County = 12 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005957
Location = CROS 8164.14.00 09/16
05/08/2018 15:03

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:04
02 Std. Gas	0.079	15:04
03 Room Air	0.000	15:05
04 Std. Gas	0.080	15:05
05 Room Air	0.000	15:05
06 Std. Gas	0.079	15:06
07 Room Air	0.000	15:06

Lot No = 17316080A2
Cyl No = 17
Exp Date = 08/05/2018
County = 12 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005957
Location = CROS 8164.14.00 09/16
05/08/2018 15:07

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:07
02 Std. Gas	0.079	15:07
03 Room Air	0.000	15:08
04 Std. Gas	0.079	15:08
05 Room Air	0.000	15:09
06 Std. Gas	0.079	15:09
07 Room Air	0.000	15:10

Lot No = 17316080A2
Cyl No = 17
Exp Date = 08/05/2018
County = 12 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005957
Location = CROS 8164.14.00 09/16
05/08/2018 15:24

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:24
02 Std. Gas	0.079	15:25
03 Room Air	0.000	15:25
04 Std. Gas	0.079	15:25
05 Room Air	0.000	15:26
06 Std. Gas	0.079	15:26
07 Room Air	0.000	15:27

Lot No = 17316080A2
Cyl No = 17
Exp Date = 08/05/2018
County = 12 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC