

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-006501

Location: GFSO

A. General Setup and Checks:

1.  Diagnostic Tests Pass and Instrument in Ready Mode  
2.  Breath tube heated  
3.  Date, time and location code (Level 2,E). Re-set if necessary.  
4.  Print test (Level 1,P). Sign and attach test record.  
5.  Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.  
a. Display: 701 psi Regulator: 700 psi

GRAND FORKS COUNTY  
SO (CORR)  
GFOCC

B. Tests (Sign and attach test records):

- Configure simulator for the following tests (Level 1,S).
1. Wet Calibration Check
- a.  Low AC (Level 1,C): Known Value  $\leq 0.03$  AC: 0.020 AC  
Sim. Ser #: DR3377 Lot #: 201610A Exp. Date: 10-4-18
- b.  High AC (Level 1,C): Known Value  $\geq 0.25$  AC: 0.300 AC  
Sim. Ser #: DR5191 Lot #: 17350 Exp. Date: 10-11-19

Configure dry gas standard for the following tests (Level 1,S).

2.  Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone  
Sim. Ser #: DR3451 Lot #: ICS #5 Exp. Date: 5-15-18  
 Display reads "Interferent Detect" CEE
3.  RFI Check (CMS Mode)  Display reads "RFI Detect"
4.  Dry Calibration Check (Level 1,C): Known Value 0.080 AC  
Gas Cyl Lot #: 19817080A9 Cyl #: 17 Exp. Date: 9-5-19
- |                      |              |        |              |        |              |
|----------------------|--------------|--------|--------------|--------|--------------|
| Test 1               | <u>0.081</u> | Test 4 | <u>0.081</u> | Test 7 | <u>0.082</u> |
| Test 2               | <u>0.082</u> | Test 5 | <u>0.082</u> | Test 8 | <u>0.081</u> |
| Test 3               | <u>0.082</u> | Test 6 | <u>0.082</u> | Test 9 | <u>0.082</u> |
| Average <u>0.082</u> |              |        |              |        |              |

C. Remarks/Maintenance: ANNUAL INSPECTION REPAIRED TEAR IN PLASTIC PROTECTIVE COVER OF BREATH HOSE

Instrument is acceptable to be used in the field.

Charles E. Edin  
Breath Analyst Signature

5-10-18  
Date

Roberta Huggins-Humms  
Reviewed by

14 May 18  
Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer           Alcohol Analyzer  
North Dakota Model 8000           SN 80-006501  
Location = GFSO                   8164.14.00 09/16  
05/10/2018                           17:27

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxy1234567890-=  
|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

abcdefghijklmnopqrstuvwxy1234567890-=  
|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

Current Instrument Setup

Data Entry Mode:           Enabled  
Start Test Sequence:       DABACABA  
Display Prelim Rslt?       Yes  
Display Third Digit?       Yes  
Inhib Printer(Y/N)?       No  
Display Volume?           No  
Disable On Memfull?       No  
# of Print Copies?         1  
Select Std (D/W/I)?       Dry  
Standard Value?           0.080  
Standard Lot #?           19817080A4  
Standard Cyl #?           17  
Standard Expiration?       09/05/2019  
Oper No?                   666666

Flow Cal. Date:           08/18/2015  
Slope                      670  
Intercept                 -629984

IR Calibration Date:       08/18/2015  
                                  3um                   9um

-----  
0th Coef(\*100):       -18086               -17820  
1st Coef(\*100):       261817               138262  
2nd Coef(\*100):       4970                   1663  
H2O adj(mg/l\*10k):   666                    427

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006501  
Location = GFSO      8164.14.00 09/16  
05/10/2018      17:41

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:41
02 Std. Sol.	0.020	17:42
03 Room Air	0.000	17:43
04 Std. Sol.	0.020	17:43
05 Room Air	0.000	17:44
06 Std. Sol.	0.020	17:45
07 Room Air	0.000	17:45

08 Sim Temp = 34.0°C

Simul Ser No = DR3377

Std Sol No = 201610A

County = 18

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

*Low AC*  
*0.020 AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006501  
Location = GFSO      8164.14.00 09/16  
05/10/2018      17:46

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:46
02 Std. Sol.	0.299	17:47
03 Room Air	0.000	17:48
04 Std. Sol.	0.300	17:49
05 Room Air	0.000	17:49
06 Std. Sol.	0.300	17:50
07 Room Air	0.000	17:51

08 Sim Temp = 34.0°C

Simul Ser No = DR5191

Std Sol No = 17350

County = 18

Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

HIGH AC  
0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006501  
Location = GFSO      8164.14.00 09/16  
05/10/2018      17:24

Test	AC	Time
01 Room Air	0.000	17:25
02 *Subject Test	INT*	17:25
03 Room Air	0.000	17:26

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 03/01/1982  
Sub Sex = Male      Weight = 120  
Test = DUI      Cit = INTERFERENT  
Dr. Lic. = ND/DIS821449  
Lot No = 19817080A4  
Cyl No = 17  
Expiration Date = 09/05/2019  
County = 18      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*Interferent Check*  
*0.10 AC + 0.05% ACETONE*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006501  
Location = GFSO                    8164.14.00 09/16  
05/10/2018                            17:31

Test	AC	Time
01 Diagnostic	OK	17:32
02 Room Air	0.000	17:32
03 *Subject Test	RFI*	17:32
04 Room Air	0.000	17:33

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 03/01/1982  
Sub Sex = Male                      Weight = 120  
Test = DUI                            Cit = RFI CHECK  
Dr. Lic. = ND/DIS821449  
Lot No = 19817080A4  
Cyl No = 17  
Expiration Date = 09/05/2019  
County = 18                          Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks:

RFI CHECK

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006501  
Location = GFSO      8164.14.00 09/16  
05/10/2018      17:33

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:34
02 Std. Gas	0.081	17:34
03 Room Air	0.000	17:35
04 Std. Gas	0.082	17:35
05 Room Air	0.000	17:36
06 Std. Gas	0.082	17:36
07 Room Air	0.000	17:37

Lot No = 19817080A4  
Cyl No = 17  
Exp Date = 09/05/2019  
County = 18      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*CALIBRATION CHECK  
0.080 AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006501  
Location = GFSO      8164.14.00 09/16  
05/10/2018      17:37

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:37
02 Std. Gas	0.081	17:38
03 Room Air	0.000	17:38
04 Std. Gas	0.082	17:39
05 Room Air	0.000	17:39
06 Std. Gas	0.082	17:39
07 Room Air	0.000	17:40

Lot No = 19817080A4  
Cyl No = 17  
Exp Date = 09/05/2019  
County = 18      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080 AC

Form 106-I8000




Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-006501  
Location = GFSO                    8164.14.00 09/16  
05/10/2018                            17:52

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:52
02 Std. Gas	0.082	17:52
03 Room Air	0.000	17:53
04 Std. Gas	0.081	17:53
05 Room Air	0.000	17:54
06 Std. Gas	0.082	17:54
07 Room Air	0.000	17:55

Lot No = 19817080A4  
Cyl No = 17  
Exp Date = 09/05/2019  
County = 18                            Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: *CALIBRATION CHECK*  
*0.080 AC*