

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-005950

Location: SHER
SHERIDAN COUNTY

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode S.O.
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 1011 psi Regulator: 1000 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a. Low AC (Level 1,C): Known Value \leq 0.03 AC: 0.020 AC
Sim. Ser #: DR3377 Lot #: 201610A Exp. Date: 10.4.18
- b. High AC (Level 1,C): Known Value \geq 0.25 AC: 0.300 AC
Sim. Ser #: DR5191 Lot #: 17350 Exp. Date: 10.11.19
CEE

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR3451 Lot #: ICS #5 Exp. Date: 5.15.19
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 34917080A3 Cyl #: 4 Exp. Date: 2/5/20

Test 1	<u>0.080</u>	Test 4	<u>0.081</u>	Test 7	<u>0.081</u>
Test 2	<u>0.080</u>	Test 5	<u>0.081</u>	Test 8	<u>0.081</u>
Test 3	<u>0.080</u>	Test 6	<u>0.080</u>	Test 9	<u>0.080</u>
Average <u>0.080</u>					

C. Remarks/Maintenance: ANNUAL INSPECTION

Instrument is acceptable to be used in the field.

Charles E. Eds
Breath Analyst Signature

5-15-2018
Date

Roberta Quequero
Reviewed by

Rob
1816 May 18
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005950
Location = SHER 8164.13.00 06/09
05/15/2018 10:10

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 34917080A3
Standard Cyl #? 4
Standard Expiration? 02/05/2020
Oper No? 666666

Flow Cal. Date: 06/20/2013
Slope 713
Intercept -723515

IR Calibration Date: 06/20/2013
3um 9um

0th Coef(*100):	-37081	-23816
1st Coef(*100):	290549	133411
2nd Coef(*100):	552	1090
H2O adj(mg/l*10k):	683	502

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005950
Location = SHER 8164.13.00 06/09
05/15/2018 10:26

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:26
02 Std. Sol.	0.018	10:27
03 Room Air	0.000	10:27
04 Std. Sol.	0.019	10:28
05 Room Air	0.000	10:29
06 Std. Sol.	0.019	10:29
07 Room Air	0.000	10:30


08 Sim Temp = 34.0°C

Simul Ser No = DR3377

Std Sol No = 201610A

County = 42

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Low AC
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005950
Location = SHER 8164.13.00 06/09
05/15/2018 10:30

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:31
02 Std. Sol.	0.294	10:32
03 Room Air	0.000	10:32
04 Std. Sol.	0.295	10:33
05 Room Air	0.000	10:34
06 Std. Sol.	0.296	10:34
07 Room Air	0.000	10:35

08 Sim Temp = 34.0°C

Simul Ser No = DR5191

Std Sol No = 17350

County = 42

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

HIGH AC
0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005950
Location = SHER 8164.13.00 06/09
05/15/2018 10:15

Test	AC	Time
01 Room Air	0.000	10:16
02 *Subject Test	INT*	10:17
03 Room Air	0.000	10:17

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 34917080A3
Cyl No = 4
Expiration Date = 02/05/2020
County = 42 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent Check
0.10 AC + 0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005950
Location = SHER 8164.13.00 06/09
05/15/2018 10:13

Test	AC	Time
01 Diagnostic	OK	10:14
02 Room Air	0.000	10:14
03 *Subject Test	RFI*	10:14
04 Room Air	0.000	10:15

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 34917080A3
Cyl No = 4
Expiration Date = 02/05/2020
County = 42 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: *RFI CHECK*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005950
Location = SHER 8164.13.00 06/09
05/15/2018 10:18

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:18
02 Std. Gas	0.080	10:18
03 Room Air	0.000	10:19
04 Std. Gas	0.080	10:19
05 Room Air	0.000	10:20
06 Std. Gas	0.080	10:20
07 Room Air	0.000	10:21

Lot No = 34917080A3
Cyl No = 4
Exp Date = 02/05/2020
County = 42 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005950
Location = SHER 8164.13.00 06/09
05/15/2018 10:21

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:22
02 Std. Gas	0.081	10:22
03 Room Air	0.000	10:22
04 Std. Gas	0.081	10:23
05 Room Air	0.000	10:23
06 Std. Gas	0.080	10:24
07 Room Air	0.000	10:24

Lot No = 34917080A3
Cyl No = 4
Exp Date = 02/05/2020
County = 42 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *CALIBRATION CHECK*
0.080 AC


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005950
Location = SHER 8164.13.00 06/09
05/15/2018 10:35

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:36
02 Std. Gas	0.081	10:36
03 Room Air	0.000	10:37
04 Std. Gas	0.081	10:37
05 Room Air	0.000	10:38
06 Std. Gas	0.080	10:38
07 Room Air	0.000	10:38

Lot No = 34917080A3
Cyl No = 4
Exp Date = 02/05/2020
County = 42 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000