

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-006500

Location: HARV
HARVEY P.D.

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 959 psi Regulator: 975 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check
 - a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: DR3377 Lot #: 201610A Exp. Date: 10.4.18
 - b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.300 AC
Sim. Ser #: DR5191 Lot #: 17350 Exp. Date: 0.300
CEE 10.11.19

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR3451 Lot #: ICS #5 Exp. Date: 5.15.19
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 03417086A3 Cyl #: 1 Exp. Date: 3/5/19

Test 1	<u>0.080</u>	Test 4	<u>0.080</u>	Test 7	<u>0.081</u>
Test 2	<u>0.080</u>	Test 5	<u>0.081</u>	Test 8	<u>0.081</u>
Test 3	<u>0.081</u>	Test 6	<u>0.081</u>	Test 9	<u>0.080</u>
Average	<u>0.081</u>				

C. Remarks/Maintenance: ANNUAL INSPECTION

Instrument is acceptable to be used in the field.

[Signature]
Breath Analyst Signature

5-15-18
Date

[Signature]
Reviewed by

16 May 18
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006500
Location = HARV 8164.13.00 06/09
05/15/2018 11:47

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxy1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 03417080A3
Standard Cyl #? 1
Standard Expiration? 03/05/2019
Oper No? 666666

Flow Cal. Date: 08/18/2015
Slope 681
Intercept -680985

IR Calibration Date: 08/18/2015
 3um 9um

0th Coef(*100): -3148 -17259
1st Coef(*100): 254475 138520
2nd Coef(*100): 1954 1329
H2O adj(mg/l*10k): 193 414

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006500
Location = HARV 8164.13.00 06/09
05/15/2018 12:02

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:03
02 Std. Sol.	0.019	12:03
03 Room Air	0.000	12:04
04 Std. Sol.	0.019	12:05
05 Room Air	0.000	12:05
06 Std. Sol.	0.019	12:06
07 Room Air	0.000	12:06

08 Sim Temp = 34.0°C

Simul Ser No = DR3377

Std Sol No = 201610A

County = 52

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Low AC

0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006500
Location = HARV 8164.13.00 06/09
05/15/2018 12:07

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:08
02 Std. Sol.	0.294	12:09
03 Room Air	0.000	12:09
04 Std. Sol.	0.297	12:10
05 Room Air	0.000	12:11
06 Std. Sol.	0.298	12:11
07 Room Air	0.000	12:12

08 Sim Temp = 34.0°C

Simul Ser No = DR5191

Std Sol No = 17350

County = 52

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

HIGH AC
0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006500
Location = HARV 8164.13.00 06/09
05/15/2018 11:52

Test	AC	Time
01 Room Air	0.000	11:53
02 *Subject Test	INT*	11:53
03 Room Air	0.000	11:54

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 03417080A3
Cyl No = 1
Expiration Date = 03/05/2019
County = 52 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent Check
0.10 AC + 0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006500
Location = HARV 8164.13.00 06/09
05/15/2018 11:48

Test	AC	Time
01 Diagnostic	OK	11:50
02 Room Air	0.000	11:50
03 *Subject Test	RFI*	11:50
04 Room Air	0.000	11:51

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 03417080A3
Cyl No = 1
Expiration Date = 03/05/2019
County = 52 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006500
Location = HARV 8164.13.00 06/09
05/15/2018 11:54

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:55
02 Std. Gas	0.080	11:55
03 Room Air	0.000	11:56
04 Std. Gas	0.080	11:56
05 Room Air	0.000	11:57
06 Std. Gas	0.081	11:57
07 Room Air	0.000	11:57

Lot No = 03417080A3
Cyl No = 1
Exp Date = 03/05/2019
County = 52 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006500
Location = HARV 8164.13.00 06/09
05/15/2018 11:58

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:58
02 Std. Gas	0.080	11:59
03 Room Air	0.000	11:59
04 Std. Gas	0.081	11:59
05 Room Air	0.000	12:00
06 Std. Gas	0.081	12:00
07 Room Air	0.000	12:01

Lot No = 03417080A3
Cyl No = 1
Exp Date = 03/05/2019
County = 52 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006500
Location = HARV 8164.13.00 06/09
05/15/2018 12:14

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:15
02 Std. Gas	0.081	12:15
03 Room Air	0.000	12:16
04 Std. Gas	0.081	12:16
05 Room Air	0.000	12:17
06 Std. Gas	0.080	12:17
07 Room Air	0.000	12:17

Lot No = 03417080A3

Cyl No = 1

Exp Date = 03/05/2019

County = 52

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION CHECK
0.080 AC

Form 106-I8000