

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-006498

Location: MINOT P.D.
(MIPD)

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 635 psi Regulator: 650 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: DR3377 Lot #: 201610A Exp. Date: 10-4-18
- b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.300 AC
Sim. Ser #: DR5191 Lot #: 17350 Exp. Date: 10-11-19

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR3451 Lot #: ICS #5 Exp. Date: 5-15-19
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 17316080A2 Cyl #: 30 Exp. Date: 8-5-18

Test 1	<u>0.081</u>	Test 4	<u>0.081</u>	Test 7	<u>0.080</u>
Test 2	<u>0.080</u>	Test 5	<u>0.080</u>	Test 8	<u>0.081</u>
Test 3	<u>0.080</u>	Test 6	<u>0.081</u>	Test 9	<u>0.080</u>
Average	<u>0.080</u>				

C. Remarks/Maintenance: ANNUAL INSPECTION

Instrument is acceptable to be used in the field.

Charles E. Eder
Breath Analyst Signature

5-17-2018
Date

Roberta Munguerra-Nimmo
Reviewed by

24 May 18
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006498
Location = MIPD 8164.14.00 09/16
05/17/2018 13:36

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:37
02 Std. Sol.	0.017	13:37
03 Room Air	0.000	13:38
04 Std. Sol.	0.018	13:39
05 Room Air	0.000	13:39
06 Std. Sol.	0.018	13:40
07 Room Air	0.000	13:40

08 Sim Temp = 34.0°C

Simul Ser No = DR3377
Std Sol No = 201610A
County = 51 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Low AC*
 0.020 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006498
Location = MIPD 8164.14.00 09/16
05/17/2018 13:47

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:48
02 Std. Sol.	0.304	13:49
03 Room Air	0.000	13:49
04 Std. Sol.	0.305	13:50
05 Room Air	0.000	13:51
06 Std. Sol.	0.305	13:51
07 Room Air	0.000	13:52


08 Sim Temp = 34.0°C

Simul Ser No = DR5191

Std Sol No = 17350

County = 51

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

HIGH AC
0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006498
Location = MIPD 8164.14.00 09/16
05/17/2018 13:17

Test	AC	Time
01 Room Air	0.000	13:18
02 *Subject Test	INT*	13:18
03 Room Air	0.000	13:19

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 17316080A2
Cyl No = 30
Expiration Date = 08/05/2018
County = 51 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent Check
0.10 AC + 0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006498
Location = MIPD 8164.14.00 09/16
05/17/2018 13:14

Test	AC	Time
01 Diagnostic	OK	13:15
02 Room Air	0.000	13:15
03 *Subject Test	RFI*	13:16
04 Room Air	0.000	13:16

*Invalid Test
 Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 17316080A2
Cyl No = 30
Expiration Date = 08/05/2018
County = 51 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006498
Location = MIPD 8164.14.00 09/16
05/17/2018 13:20

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:20
02 Std. Gas	0.081	13:20
03 Room Air	0.000	13:21
04 Std. Gas	0.080	13:21
05 Room Air	0.000	13:22
06 Std. Gas	0.080	13:22
07 Room Air	0.000	13:23

Lot No = 17316080A2
Cyl No = 30
Exp Date = 08/05/2018
County = 51 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *CALIBRATION CHECK*
0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006498
Location = MIPD 8164.14.00 09/16
05/17/2018 13:24

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:24
02 Std. Gas	0.081	13:25
03 Room Air	0.000	13:25
04 Std. Gas	0.080	13:25
05 Room Air	0.000	13:26
06 Std. Gas	0.081	13:26
07 Room Air	0.000	13:27

Lot No = 17316080A2
Cyl No = 30
Exp Date = 08/05/2018
County = 51 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006498
Location = MIPD 8164.14.00 09/16
05/17/2018 13:54

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:55
02 Std. Gas	0.080	13:55
03 Room Air	0.000	13:55
04 Std. Gas	0.081	13:56
05 Room Air	0.000	13:56
06 Std. Gas	0.080	13:57
07 Room Air	0.000	13:57

Lot No = 17316080A2

Cyl No = 30

Exp Date = 08/05/2018

County = 51

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION CHECK
0.080 AC

Form 106-I8000