

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-005938

ASHL  
Location: McIntosh Casco.

A. General Setup and Checks:

1.  Diagnostic Tests Pass and Instrument in Ready Mode
2.  Breath tube heated
3.  Date, time and location code (Level 2,E). Re-set if necessary.
4.  Print test (Level 1,P). Sign and attach test record.
5.  Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
  - a. Display: 951 psi Regulator: 975 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a.  Low AC (Level 1,C): Known Value  $\leq 0.03$  AC: 0.020 AC  
Sim. Ser #: DR3377 Lot #: 201610A Exp. Date: 4 Oct 18
- b.  High AC (Level 1,C): Known Value  $\geq 0.25$  AC: 0.300 AC  
Sim. Ser #: DR5911 Lot #: 17350 Exp. Date: 11 Oct 19

Configure dry gas standard for the following tests (Level 1,S).

2.  Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone  
Sim. Ser #: DR3451 Lot #: 1C55 Exp. Date: 15 May 19  
 Display reads "Interferent Detect"
3.  RFI Check (CMS Mode)  Display reads "RFI Detect"
4.  Dry Calibration Check (Level 1,C): Known Value 0.080 AC  
Gas Cyl Lot #: 34917080A3 Cyl #: 33 Exp. Date: 5 Feb 20

Test 1 <u>0.079</u>	Test 4 <u>0.078</u>	Test 7 <u>0.078</u>
Test 2 <u>0.079</u>	Test 5 <u>0.079</u>	Test 8 <u>0.079</u>
Test 3 <u>0.078</u>	Test 6 <u>0.077</u>	Test 9 <u>0.079</u>
Average <u>0.078</u>		

C. Remarks/Maintenance: Annual inspection. Completed tare of tank sensor. Replaced o-ring on simulator return.

Instrument is acceptable to be used in the field.

Roberta Sueque-Nimmo  
Breath Analyst Signature

25 May 18  
Date

Charles E. Eden  
Reviewed by

25 MAY 18  
Date



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005938  
Location = ASHL      8164.14.00 09/16  
05/25/2018      11:18

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:18
02 Std. Sol.	0.019	11:19
03 Room Air	0.000	11:20
04 Std. Sol.	0.019	11:20
05 Room Air	0.000	11:21
06 Std. Sol.	0.019	11:22
07 Room Air	0.000	11:22


08 Sim Temp = 34.0°C

Simul Ser No = DR3377

Std Sol No = 201610A

County = 26

Oper No. = 888888



Operator Signature  
ROBERTA GRIEGER-NIMMO

Remarks: *Low AC*  
*0.020 AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005938  
Location = ASHL      8164.14.00 09/16  
05/25/2018      11:10

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:11
02 Std. Sol.	0.294	11:12
03 Room Air	0.000	11:12
04 Std. Sol.	0.295	11:13
05 Room Air	0.000	11:13
06 Std. Sol.	0.295	11:14
07 Room Air	0.000	11:15

08 Sim Temp = 34.0°C

Simul Ser No = DR5191

Std Sol No = 17350

County = 08

Oper No. = 888888

Roberta Grieger-Nimmo

Operator Signature  
ROBERTA GRIEGER-NIMMO

Remarks: High AC  
0.300 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005938  
Location = ASHL      8164.14.00 09/16  
05/25/2018      10:49

Test	AC	Time
01 Room Air	0.000	10:51
02 *Subject Test	INT*	10:51
03 Room Air	0.000	10:52

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = INTERFERENT  
Dr. Lic. = ND/DIS821456  
Lot No = 34917080A3  
Cyl No = 33  
Expiration Date = 02/05/2020  
County = 26      Oper No. = 888888

  
Operator Signature  
ROBERTA GRIEGER-NIMMO

Remarks: Interferent Check  
0.10 AC + 0.05% Acetone

Form 106-I8000



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005938  
Location = ASHL                    8164.14.00 09/16  
05/25/2018    10:57

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:57
02 Std. Gas	0.079	10:58
03 Room Air	0.000	10:58
04 Std. Gas	0.079	10:59
05 Room Air	0.000	10:59
06 Std. Gas	0.078	10:59
07 Room Air	0.000	11:00

Lot No = 34917080A3  
Cyl No = 33  
Exp Date = 02/05/2020  
County = 26    Oper No. = 888888



Operator Signature  
ROBERTA GRIEGER-NIMMO

Remarks: Calibration Check  
0.080AC


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005938  
Location = ASHL                    8164.14.00 09/16  
05/25/2018                            11:00

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:01
02 Std. Gas	0.078	11:01
03 Room Air	0.000	11:02
04 Std. Gas	0.079	11:02
05 Room Air	0.000	11:03
06 Std. Gas	0.077	11:03
07 Room Air	0.000	11:03

Lot No = 34917080A3  
Cyl No = 33  
Exp Date = 02/05/2020  
County = 26                            Oper No. = 888888

  
Operator Signature  
ROBERTA GRIEGER-NIMMO

Remarks: *Calibration Check*  
*0.080 AC*



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005938  
Location = ASHL      8164.14.00 09/16  
05/25/2018      11:04

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:05
02 Std. Gas	0.078	11:05
03 Room Air	0.000	11:06
04 Std. Gas	0.079	11:06
05 Room Air	0.000	11:07
06 Std. Gas	0.079	11:07
07 Room Air	0.000	11:07

Lot No = 34917080A3  
Cyl No = 33  
Exp Date = 02/05/2020  
County = 26      Oper No. = 888888

*Roberta Grieger-Nimmo*

Operator Signature  
ROBERTA GRIEGER-NIMMO

Remarks: *Calibration Check*  
*0.080 AC*