

CHAIN OF CUSTODY
BREATH ALCOHOL INSTRUMENT

Instrument: 8000 Serial Number: 80-005949

Agency being assigned instrument: DUMP

Date 28 Jun 18

Transferred from:

Transferred to:

Name (print) Roberta Grieger-Nimmö Name (print) DUMP

Signature Roberta Grieger-Nimmö Signature NA

Agency Crime Lab Agency NA

Reason for transfer/remarks:

Instrument is unable to turn on due to power supply failure. It is not cost effective to repair instrument. Useable parts will be salvaged to repair other instruments.

Or Shipped Via:

USPS _____ UPS _____ FedEx _____ Other _____

Date Sending Instrument: _____ / _____ / _____ Air Bill No. _____