



# ETHANOL GAS STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program  
SFN 59282 (10-2013)

|   |   |                                     |
|---|---|-------------------------------------|
| Chemical Test Operator Name (Print) <i>Amber Johnson 021112</i> |   |                                     |
| Location <i>Fort Yates</i>                                      | Intoxilyzer® Serial Number <i>80-006492</i> |                                     |
| Gas Lot Number <i>349 170 80 A3</i>                             | Gas Cylinder Number <i>29</i>               | Gas Expiration Date <i>2-5-2020</i> |

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).  
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
  - A. *0.078* AC
  - B. *0.079* AC
  - C. *0.079* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
  - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
  - B. ACA Test Record.

|   |                     |
|---|---------------------|
| Chemical Test Operator Signature <i>Amber A. Johnson</i>            | Date <i>8/9/18</i>  |
| Reviewed By (Crime Laboratory Use Only) <i>Roberta Huergo-Nummo</i> | Date <i>8/10/18</i> |


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006492  
Location = FYPD      8164.14.00 09/16  
08/09/2018      11:03

DRY CAL CHECK

| Test        | AC    | Time  |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 11:03 |
| 02 Std. Gas | 0.078 | 11:04 |
| 03 Room Air | 0.000 | 11:04 |
| 04 Std. Gas | 0.079 | 11:05 |
| 05 Room Air | 0.000 | 11:05 |
| 06 Std. Gas | 0.079 | 11:06 |
| 07 Room Air | 0.000 | 11:06 |

Lot No = 34917080A3  
Cyl No = 29  
Exp Date = 02/05/2020  
County = 61      Oper No. = 021112

  
Operator Signature  
AMBER JOHNSON

Remarks:  
Gas install - Checks ok.