



ETHANOL GAS STANDARD CYLINDER REPORT
 OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION
 Toxicology Section/Breath Alcohol Program
 SFN 59282 (10-2013)

Chemical Test Operator Name (Print) <i>Christopher Adam Secrest</i>		
Location <i>205 Main STE Mohall, NP 58761</i>	Intoxilyzer® Serial Number <i>80-006505</i>	
Gas Lot Number <i>3491708A3</i> <i>34917080A3</i>	Gas Cylinder Number <i>031</i>	Gas Expiration Date <i>02/05/2020</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 A. 0.080 AC
 B. 0.079 AC
 C. 0.079 AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>[Signature]</i>	Date <i>08/29/18</i>
Reviewed By (Crime Laboratory Use Only) <i>Robert [Signature]</i>	Date <i>08/30/18</i>

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006505
Location = MOHL 8164.14.00 09/16
08/29/2018 16:22

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:23
02 Std. Gas	0.080	16:23
03 Room Air	0.000	16:24
04 Std. Gas	0.079	16:24
05 Room Air	0.000	16:25
06 Std. Gas	0.079	16:25
07 Room Air	0.000	16:26


Lot No = 34917080A3

Cyl No = 31

Exp Date = 02/05/2020

County = 38

Oper No. = 130839


Operator Signature
CHRISTOPHER SECREST

Remarks:

Form 106-I8000