



**INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT**  
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM  
 SFN 59281 (06/2018)

Serial Number 80-005955	Instrument Location Dunn County Sheriff's Office, 205 Owens St Manning, ND, 58446
Reason for Install/Repair <input type="checkbox"/> Install After Receiving From Crime Laboratory <input checked="" type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
  - A. Print Test (Level 1, Function P).
  - B. ACA Test (Level 1, Function C).
  - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): None
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature 	Date 09/06/2018
-------------------------------	--------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By 	Certified Date 09/06/2018
---------------------------	------------------------------

I certify that the attached is a true and correct copy of the intoxilyzer test record.  
 Print Name/Badge #: HUNTER RHOADES /3261  
 Signature:   
 Date: 9/6/18 Dept: DUNN COUNTY S.O.

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005955  
Location = DUNN                    8164.14.00 09/16  
09/05/2018                            17:40

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxy1234567890-|=|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&\*()\_+?

abcdefghijklmnopqrstuvwxy1234567890-|=|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&\*()\_+?

Current Instrument Setup

Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?          1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            03417080A3  
Standard Cyl #?            031  
Standard Expiration?       03/05/2019  
Oper No?                    130635

Flow Cal. Date:            06/21/2013  
Slope                      757  
Intercept                  -837385

IR Calibration Date:       06/21/2013  
                                    3um                      9um

	3um	9um
0th Coef(*100):	-19897	-22878
1st Coef(*100):	268381	131979
2nd Coef(*100):	5157	1473
H2O adj(mg/l*10k):	717	600

\*\*\*\*\* Printer Test End \*\*\*\*\*

  
\_\_\_\_\_  
Operator Signature  
HUNTER RHOADES

Remarks:

Intoxilyzer Test Record and Checklist  
 NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
 North Dakota Model 8000      SN 80-005955  
 Location = DUNN      8164.14.00 09/16  
 09/05/2018      17:43

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:44
02 Std. Gas	0.082	17:44
03 Room Air	0.000	17:45
04 Std. Gas	0.081	17:45
05 Room Air	0.000	17:46
06 Std. Gas	0.082	17:46
07 Room Air	0.000	17:47

Lot No = 03417080A3  
 Cyl No = 031  
 Exp Date = 03/05/2019  
 County = 13      Oper No. = 130635



Operator Signature  
 HUNTER RHOADES

Remarks:

Form 106-I8000

I certify that the attached is a true and  
 correct copy of the intoxilyzer test record.

Print Name/Badge #: GRL HUNTER RHOADES #3261

Signature: 

Date: 9/5/18      Dept: DUNN COUNTY S.O.


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer                      Alcohol Analyzer  
North Dakota Model 8000                      SN 80-005955  
Location = DUNN                              8164.14.00 09/16  
09/06/2018    09:24

Test	AC	Time
01 Room Air	RFI*	09:25
02 Room Air	RFI*	09:25


\*Invalid Test  
Inhibited - RFI

Sub Name = TEST, RFI NA  
Sub DOB = 01/01/1920  
Sub Sex = Male                              Weight = 900  
Test = OTH                                      Cit = NA  
Dr. Lic. = ND/000000000  
Lot No = 03417080A3  
Cyl No = 031  
Expiration Date = 03/05/2019  
County = 13                              Oper No. = 130635

  
Operator Signature  
HUNTER RHOADES

Remarks: RFI TEST

Form 106-I8000

I certify that the attached is a true and correct copy of the intoxilyzer test record.  
Print Name/Badge #: HUNTER RHOADES/3261  
Signature:   
Date: 9/6/18 Dept: DUNN COUNTY S.O.