



ETHANOL GAS STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (10-2013)

Chemical Test Operator Name (Print) <i>Jonathan Lampi</i>		
Location <i>GFPD</i>	Intoxilyzer® Serial Number <i>80-006489</i>	
Gas Lot Number <i>34917080 A3</i>	Gas Cylinder Number <i>12</i>	Gas Expiration Date <i>02-05-20</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. *0.079* AC
 - B. *0.080* AC
 - C. *0.078* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>Jon Lampi</i>	Date <i>09-12-18</i>
Reviewed By (Crime Laboratory Use Only) <i>Roberta Plunger-Nummo</i>	Date <i>09-12-18</i>


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006489
Location = GFPD 8164.14.00 09/16
09/12/2018 12:32

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:33
02 Std. Gas	0.079	12:33
03 Room Air	0.000	12:34
04 Std. Gas	0.080	12:34
05 Room Air	0.000	12:35
06 Std. Gas	0.078	12:35
07 Room Air	0.000	12:36

Lot No = 34917080A3
Cyl No = 12
Exp Date = 02/05/2020
County = 18 Oper No. = 039947


Operator Signature
JONATHAN LAMPI

Remarks:
ACA For Gas Standard Cylinder report