



ETHANOL GAS STANDARD CYLINDER REPORT
NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Jonathan Davis</i>		
Location <i>WCPD / MCLF</i>	Intoxilyzer® Serial Number <i>80-005958</i>	
Gas Lot Number <i>17618080A1</i>	Gas Cylinder Number <i>29</i>	Gas Expiration Date <i>8-5-20</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. *0.026* AC
 - B. *0.027* AC
 - C. *0.028* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>[Signature]</i>	Date <i>3-26-19</i>
Reviewed By (Crime Laboratory Use Only) <i>[Signature]</i>	Date <i>3-29-19</i>


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005958
Location = WCPD 8164.14.00 09/16
03/26/2019 22:54

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	22:54
02 Std. Gas	0.076	22:55
03 Room Air	0.000	22:55
04 Std. Gas	0.077	22:55
05 Room Air	0.000	22:56
06 Std. Gas	0.078	22:56
07 Room Air	0.000	22:57

Lot No = 17618080A1
Cyl No = 29
Exp Date = 08/05/2020
County = 27 Oper No. = 131748


Operator Signature
JONATHAN DAVIS

Remarks: *Valid*

Form 106-I8000