



ETHANOL GAS STANDARD CYLINDER REPORT

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Kelly, Charles L</i>		
Location <i>Stutsman LEC Jail Intake</i>	Intoxilyzer® Serial Number <i>80-006499</i>	
Gas Lot Number <i>13518080 AG</i>	Gas Cylinder Number <i>36</i>	Gas Expiration Date <i>8-5-20</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. *0.080* AC
 - B. *0.081* AC
 - C. *0.081* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>[Signature]</i> #340	Date <i>3-27-19</i>
Reviewed By (Crime Laboratory Use Only) <i>[Signature]</i>	Date <i>4.8.2019</i> CEE

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006499
Location = JAME 8164.14.00 09/16
03/27/2019 14:33

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:33
02 Std. Gas	0.082	14:34
03 Room Air	0.000	14:34
04 Std. Gas	0.081	14:35
05 Room Air	RFI*	14:35
06 Room Air	0.000	14:35

*Invalid Test
Inhibited - RFI

Lot No = 13518080A6
Cyl No = 36
Exp Date = 08/05/2020
County = 47 Oper No. = 131332

 #340

Operator Signature
CHARLES KELLY

Remarks:

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006499
Location = JAME 8164.14.00 09/16
03/27/2019 14:36

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:37
02 Std. Gas	0.080	14:37
03 Room Air	0.000	14:38
04 Std. Gas	0.081	14:38
05 Room Air	0.000	14:38
06 Std. Gas	0.081	14:39
07 Room Air	0.000	14:39

Lot No = 13518080A6
Cyl No = 36
Exp Date = 08/05/2020
County = 47 Oper No. = 131332

 #340

Operator Signature
CHARLES KELLY

Remarks:

Form 106-I8000