

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006491
Location = BEPD 8164.14.00 09/16
04/18/2019 15:28

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:28
02 Std. Sol.	0.020	15:29
03 Room Air	0.000	15:30
04 Std. Sol.	0.020	15:30
05 Room Air	0.000	15:31
06 Std. Sol.	0.020	15:32
07 Room Air	0.000	15:32

08 Sim Temp = 34.0°C

Simul Ser No = DR3375
Std Sol No = 17160
County = 29 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *LOW AC*
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006491
Location = BEPD 8164.14.00 09/16
04/18/2019 15:34

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:35
02 Std. Sol.	0.302	15:36
03 Room Air	0.000	15:36
04 Std. Sol.	0.304	15:37
05 Room Air	0.000	15:37
06 Std. Sol.	0.303	15:38
07 Room Air	0.000	15:39


08 Sim Temp = 34.0°C

Simul Ser No = DR7111

Std Sol No = 17350

County = 29

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

HIGH AC
0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006491
Location = BEPD 8164.14.00 09/16
04/18/2019 15:16

Test	AC	Time
01 Room Air	0.000	15:17
02 *Subject Test	INT*	15:17
03 Room Air	0.000	15:18

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 34917080A3
Cyl No = 21
Expiration Date = 02/05/2020
County = 29 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent Check
0.10AC + 0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006491
Location = BEPD 8164.14.00 09/16
04/18/2019 15:19

Test	AC	Time
01 Diagnostic	OK	15:20
02 Room Air	0.000	15:20
03 *Subject Test	RFI*	15:20
04 Room Air	0.000	15:21

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 34917080A3
Cyl No = 21
Expiration Date = 02/05/2020
County = 29 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006491
Location = BEPD 8164.14.00 09/16
04/18/2019 15:11

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:12
02 Std. Gas	0.082	15:13
03 Room Air	0.000	15:13
04 Std. Gas	0.081	15:14
05 Room Air	0.000	15:14
06 Std. Gas	0.082	15:15
07 Room Air	0.000	15:15

Lot No = 34917080A3
Cyl No = 21
Exp Date = 02/05/2020
County = 29 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006491
Location = BEPD 8164.14.00 09/16
04/18/2019 15:22

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:22
02 Std. Gas	0.081	15:23
03 Room Air	0.000	15:23
04 Std. Gas	0.081	15:24
05 Room Air	0.000	15:24
06 Std. Gas	0.081	15:25
07 Room Air	0.000	15:25

Lot No = 34917080A3
Cyl No = 21
Exp Date = 02/05/2020
County = 29 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *CALIBRATION CHECK*
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006491
Location = BEPD 8164.14.00 09/16
04/18/2019 15:41

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:42
02 Std. Gas	0.080	15:42
03 Room Air	0.000	15:43
04 Std. Gas	0.081	15:43
05 Room Air	0.000	15:44
06 Std. Gas	0.081	15:44
07 Room Air	0.000	15:44

Lot No = 34917080A3
Cyl No = 21
Exp Date = 02/05/2020
County = 29 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080AC
Form 106-I8000