

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-004942

Location: MERCER COUNTY
S.O. (STAT)

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 961 psi Regulator: 1000 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: DR3375 Lot #: 17160 Exp. Date: 6.8.19
- b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.300 AC
Sim. Ser #: DR7111 Lot #: 17350 Exp. Date: 10.11.19

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR7089 Lot #: ICS 5 Exp. Date: NA
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 13518080AG Cyl #: 35 Exp. Date: 8.5.20

Test 1 <u>0.083</u>	Test 4 <u>0.082</u>	Test 7 <u>0.082</u>
Test 2 <u>0.081</u>	Test 5 <u>0.082</u>	Test 8 <u>0.082</u>
Test 3 <u>0.081</u>	Test 6 <u>0.081</u>	Test 9 <u>0.081</u>
Average <u>0.082</u>		

C. Remarks/Maintenance: ANNUAL INSPECTION

Instrument is acceptable to be used in the field.

Charles E. E...
Breath Analyst Signature

4.23.2019
Date

NA
Reviewed by

NA
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004942
Location = STAT 8164.14.00 09/16
04/23/2019 15:33

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890- = |
ABCDEFGHIJKLMN OPQRSTUVWXYZ!@# \$%^&* () _ + ?

abcdefghijklmnopqrstuvwxyz1234567890- = |
ABCDEFGHIJKLMN OPQRSTUVWXYZ!@# \$%^&* () _ + ?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer (Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 13518080A6
Standard Cyl #? 035
Standard Expiration? 08/05/2020
Oper No? 666666

Flow Cal. Date: 05/25/2018
Slope 626
Intercept -548004

IR Calibration Date: 05/25/2018
 3um 9um

0th Coef (*100): -10942 -13713
1st Coef (*100): 255654 132286
2nd Coef (*100): 3705 1469
H2O adj (mg/l*10k): 481 462

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004942
Location = STAT 8164.14.00 09/16
04/23/2019 15:52

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:52
02 Std. Sol.	0.021	15:53
03 Room Air	0.000	15:54
04 Std. Sol.	0.021	15:55
05 Room Air	0.000	15:55
06 Std. Sol.	0.021	15:56
07 Room Air	0.000	15:56

08 Sim Temp = 34.0°C

Simul Ser No = DR3375

Std Sol No = 17160

County = 29

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Low AC
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004942
Location = STAT 8164.14.00 09/16
04/23/2019 15:58

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:59
02 Std. Sol.	0.305	16:00
03 Room Air	0.000	16:00
04 Std. Sol.	0.306	16:01
05 Room Air	0.000	16:01
06 Std. Sol.	0.307	16:02
07 Room Air	0.000	16:03

08 Sim Temp = 34.0°C

Simul Ser No = DR7111

Std Sol No = 17350

County = 29

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

HIGH AC
0.300 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004942
Location = STAT 8164.14.00 09/16
04/23/2019 15:44

Test	AC	Time
01 Room Air	0.000	15:45
02 *Subject Test	INT*	15:46
03 Room Air	0.000	15:46

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 13518080A6
Cyl No = 035
Expiration Date = 08/05/2020
County = 29 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent Check
0.10 AC + 0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004942
Location = STAT 8164.14.00 09/16
04/23/2019 15:41

Test	AC	Time
01 Diagnostic	OK	15:42
02 Room Air	0.000	15:43
03 *Subject Test	RFI*	15:43
04 Room Air	0.000	15:43

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 13518080A6
Cyl No = 035
Expiration Date = 08/05/2020
County = 29 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004942
Location = STAT 8164.14.00 09/16
04/23/2019 15:36

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:37
02 Std. Gas	0.083	15:37
03 Room Air	0.000	15:38
04 Std. Gas	0.081	15:38
05 Room Air	0.000	15:38
06 Std. Gas	0.081	15:39
07 Room Air	0.000	15:39

Lot No = 13518080A6
Cyl No = 035
Exp Date = 08/05/2020
County = 29 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION CHECK
0.080 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004942
Location = STAT 8164.14.00 09/16
04/23/2019 15:48

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:49
02 Std. Gas	0.082	15:49
03 Room Air	0.000	15:50
04 Std. Gas	0.082	15:50
05 Room Air	0.000	15:50
06 Std. Gas	0.081	15:51
07 Room Air	0.000	15:51

Lot No = 13518080A6
Cyl No = 035
Exp Date = 08/05/2020
County = 29 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004942
Location = STAT 8164.14.00 09/16
04/23/2019 16:05

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:06
02 Std. Gas	0.082	16:06
03 Room Air	0.000	16:06
04 Std. Gas	0.082	16:07
05 Room Air	0.000	16:07
06 Std. Gas	0.081	16:08
07 Room Air	0.000	16:08

Lot No = 13518080A6
Cyl No = 035
Exp Date = 08/05/2020
County = 29 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION CHECK

0.080 AC

Form 106-I8000