

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-005956

Location: OLIVER COUNTY
S.O. (OLIV)

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 955 psi Regulator: 1000 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: DR3375 Lot #: 17160 Exp. Date: 6-8-19
- b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.300 AC
Sim. Ser #: DR7111 Lot #: 17350 Exp. Date: 10-11-19

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR7089 Lot #: ICS 5 Exp. Date: NA
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 39917080A3 Cyl #: 34 Exp. Date: 2/5/20
Test 1 0.078 Test 4 0.077 Test 7 0.077
Test 2 0.077 Test 5 0.077 Test 8 0.077
Test 3 0.078 Test 6 0.078 Test 9 0.077
Average 0.077

C. Remarks/Maintenance: ANNUAL INSPECTION

Instrument is acceptable to be used in the field.

Charles E. Eden
Breath Analyst Signature
NA

CEE

4-230-19
Date
NA

Reviewed by _____

Date _____

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005956
Location = OLIV 8164.14.00 09/16
04/30/2019 13:30

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-|=|
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-|=|
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 34917080A3
Standard Cyl #? 34
Standard Expiration? 02/05/2020
Oper No? 666666

Flow Cal. Date: 06/21/2013
Slope 706
Intercept -720901

IR Calibration Date: 04/02/2015
 3um 9um

0th Coef(*100): -23934 -22166
1st Coef(*100): 279293 133410
2nd Coef(*100): 3061 1559
H2O adj(mg/l*10k): 483 423

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005956
Location = OLIV 8164.14.00 09/16
04/30/2019 13:51

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:51
02 Std. Sol.	0.019	13:52
03 Room Air	0.000	13:53
04 Std. Sol.	0.019	13:54
05 Room Air	0.000	13:54
06 Std. Sol.	0.020	13:55
07 Room Air	0.000	13:55

08 Sim Temp = 34.0°C

Simul Ser No = DR3375
Std Sol No = 17160
County = 33 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Low AC*
 0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005956
Location = OLIV 8164.14.00 09/16
04/30/2019 13:57

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:57
02 Std. Sol.	0.303	13:58
03 Room Air	0.000	13:59
04 Std. Sol.	0.304	14:00
05 Room Air	0.000	14:00
06 Std. Sol.	0.303	14:01
07 Room Air	0.000	14:01

08 Sim Temp = 34.0°C

Simul Ser No = DR7111
Std Sol No = 17350
County = 33 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *HIGH AC*
0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005956
Location = OLIV 8164.14.00 09/16
04/30/2019 13:46

Test	AC	Time
01 Room Air	0.000	13:47
02 *Subject Test	INT*	13:48
03 Room Air	0.000	13:49

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 34917080A3
Cyl No = 34
Expiration Date = 02/05/2020
County = 33 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent Check
0.10AC + 0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005956
Location = OLIV 8164.14.00 09/16
04/30/2019 13:38

Test	AC	Time
01 Diagnostic	OK	13:39
02 Room Air	0.000	13:40
03 *Subject Test	RFI*	13:40
04 Room Air	0.000	13:40

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 34917080A3
Cyl No = 34
Expiration Date = 02/05/2020
County = 33 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005956
Location = OLIV 8164.14.00 09/16
04/30/2019 13:32

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:33
02 Std. Gas	0.078	13:34
03 Room Air	0.000	13:34
04 Std. Gas	0.077	13:35
05 Room Air	0.000	13:35
06 Std. Gas	0.078	13:35
07 Room Air	0.000	13:36

Lot No = 34917080A3
Cyl No = 34
Exp Date = 02/05/2020
County = 33 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005956
Location = OLIV 8164.14.00 09/16
04/30/2019 13:41

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:42
02 Std. Gas	0.077	13:42
03 Room Air	0.000	13:43
04 Std. Gas	0.077	13:43
05 Room Air	0.000	13:44
06 Std. Gas	0.078	13:44
07 Room Air	0.000	13:45

Lot No = 34917080A3
Cyl No = 34
Exp Date = 02/05/2020
County = 33 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005956
Location = OLIV 8164.14.00 09/16
04/30/2019 14:04

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:05
02 Std. Gas	0.077	14:05
03 Room Air	0.000	14:06
04 Std. Gas	0.077	14:06
05 Room Air	0.000	14:06
06 Std. Gas	0.077	14:07
07 Room Air	0.000	14:07

Lot No = 34917080A3
Cyl No = 34
Exp Date = 02/05/2020
County = 33 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000