

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-005945

Location: WATF  
MCKENZIE COUNTY  
LEC

A. General Setup and Checks:

1.  Diagnostic Tests Pass and Instrument in Ready Mode
2.  Breath tube heated
3.  Date, time and location code (Level 2,E). Re-set if necessary.
4.  Print test (Level 1,P). Sign and attach test record.
5.  Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
  - a. Display: 393 psi Regulator: 400 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a.  Low AC (Level 1,C): Known Value  $\leq 0.03$  AC: 0.020 AC  
Sim. Ser #: DR3375 Lot #: 17160 Exp. Date: 6.8.19
- b.  High AC (Level 1,C): Known Value  $\geq 0.25$  AC: 0.300 AC  
Sim. Ser #: DR7111 Lot #: 17350 Exp. Date: 10.11.19

Configure dry gas standard for the following tests (Level 1,S).

2.  Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone  
Sim. Ser #: DR7089 Lot #: ICS 5 Exp. Date: NA  
 Display reads "Interferent Detect"
3.  RFI Check (CMS Mode)  Display reads "RFI Detect"
4.  Dry Calibration Check (Level 1,C): Known Value 0.080 AC  
Gas Cyl Lot #: 198170804 Cyl #: 18 Exp. Date: 9.5.2019

Test 1 <u>0.080</u>	Test 4 <u>0.078</u>	Test 7 <u>0.078</u>
Test 2 <u>0.078</u>	Test 5 <u>0.079</u>	Test 8 <u>0.079</u>
Test 3 <u>0.079</u>	Test 6 <u>0.079</u>	Test 9 <u>0.078</u>
Average <u>0.079</u>		

C. Remarks/Maintenance: ANNUAL INSPECTION

Instrument is acceptable to be used in the field.

Charles E. Ed  
Breath Analyst Signature

5.3.2019  
Date

NA

NA

Reviewed by

Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005945  
Location = WATF                8164.14.00 09/16  
05/03/2019                      08:58

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890- = |  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890- = |  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

Current Instrument Setup

Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?         1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            19817080A4  
Standard Cyl #?            018  
Standard Expiration?       09/05/2019  
Oper No?                    666666

Flow Cal. Date:            06/20/2013  
Slope                      637  
Intercept                  -566407

IR Calibration Date:       05/22/2015

	3um	9um
0th Coef(*100):	-21213	-23093
1st Coef(*100):	279171	139292
2nd Coef(*100):	2898	1291
H2O adj(mg/l*10k):	332	319

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
CHARLES EDER

Remarks:

PRINT TEST

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005945  
Location = WATF      8164.14.00 09/16  
05/03/2019      09:18

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:18
02 Std. Sol.	0.020	09:19
03 Room Air	0.000	09:20
04 Std. Sol.	0.019	09:20
05 Room Air	0.000	09:21
06 Std. Sol.	0.020	09:22
07 Room Air	0.000	09:22

08 Sim Temp = 34.0°C

Simul Ser No = DR3375

Std Sol No = 17160

County = 27

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

*LOW AC*

*0.020 AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005945  
Location = WATF      8164.14.00 09/16  
05/03/2019      09:25

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:26
02 Std. Sol.	0.307	09:26
03 Room Air	0.000	09:27
04 Std. Sol.	0.307	09:28
05 Room Air	0.000	09:28
06 Std. Sol.	0.308	09:29
07 Room Air	0.000	09:30

08 Sim Temp = 34.0°C

Simul Ser No = DR7111  
Std Sol No = 17350  
County = 27      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

H76A AC  
0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005945  
Location = WATF      8164.14.00 09/16  
05/03/2019      09:05

Test	AC	Time
01 Room Air	0.000	09:06
02 *Subject Test	INT*	09:06
03 Room Air	0.000	09:07

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = INTERFERENT  
Dr. Lic. = ND/DIS821456  
Lot No = 19817080A4  
Cyl No = 018  
Expiration Date = 09/05/2019  
County = 27      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: *Interferent Check*  
*0.10 AC + 0.05% ACETONE*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005945  
Location = WATF                    8164.14.00 09/16  
05/03/2019                            09:13

Test	AC	Time
01 Diagnostic	OK	09:14
02 Room Air	0.000	09:15
03 *Subject Test	RFI*	09:15
04 Room Air	0.000	09:15

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = RFI CHECK  
Dr. Lic. = ND/DIS821456  
Lot No = 19817080A4  
Cyl No = 018  
Expiration Date = 09/05/2019  
County = 27                            Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks: *RFI CHECK*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005945  
Location = WATF                    8164.14.00 09/16  
05/03/2019                            08:54

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:55
02 Std. Gas	0.080	08:55
03 Room Air	0.000	08:56
04 Std. Gas	0.078	08:56
05 Room Air	0.000	08:57
06 Std. Gas	0.079	08:57
07 Room Air	0.000	08:58

Lot No = 19817080A4  
Cyl No = 018  
Exp Date = 09/05/2019  
County = 27                            Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: *CALIBRATION CHECK*  
*0.080 AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005945  
Location = WATF      8164.14.00 09/16  
05/03/2019      09:08

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:08
02 Std. Gas	0.078	09:09
03 Room Air	0.000	09:09
04 Std. Gas	0.079	09:10
05 Room Air	0.000	09:10
06 Std. Gas	0.079	09:10
07 Room Air	0.000	09:11

Lot No = 19817080A4  
Cyl No = 018  
Exp Date = 09/05/2019  
County = 27      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080 AC

Form 106-I8000



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005945  
Location = WATF      8164.14.00 09/16  
05/03/2019      09:31

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:32
02 Std. Gas	0.078	09:32
03 Room Air	0.000	09:33
04 Std. Gas	0.079	09:33
05 Room Air	0.000	09:34
06 Std. Gas	0.078	09:34
07 Room Air	0.000	09:35

Lot No = 19817080A4  
Cyl No = 18  
Exp Date = 09/05/2019  
County = 27      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: **CALIBRATION CHECK**  
**0.080 AC**

Form 106-I8000