

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-006500

Location: HARVEY P.D.  
(HARV)

A. General Setup and Checks:

1.  Diagnostic Tests Pass and Instrument in Ready Mode
2.  Breath tube heated
3.  Date, time and location code (Level 2,E). Re-set if necessary.
4.  Print test (Level 1,P). Sign and attach test record.
5.  Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
  - a. Display: 1004 psi Regulator: 1000 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a.  Low AC (Level 1,C): Known Value  $\leq 0.03$  AC: 0.020 AC  
Sim. Ser #: DR3375 Lot #: 17160 Exp. Date: 6.8.19
- b.  High AC (Level 1,C): Known Value  $\geq 0.25$  AC: 0.300 AC  
Sim. Ser #: DR7111 Lot #: 17350 Exp. Date: 10.11.19

Configure dry gas standard for the following tests (Level 1,S).

2.  Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone  
Sim. Ser #: DR7089 Lot #: ICS 5 Exp. Date: NA

Display reads "Interferent Detect"

3.  RFI Check (CMS Mode)  Display reads "RFI Detect"

4.  Dry Calibration Check (Level 1,C): Known Value 0.080 AC  
Gas Cyl Lot #: 34418080A2 Cyl #: 61 Exp. Date: 2.5.21

Test 1	<u>0.080</u>	Test 4	<u>0.081</u>	Test 7	<u>0.081</u>
Test 2	<u>0.081</u>	Test 5	<u>0.080</u>	Test 8	<u>0.081</u>
Test 3	<u>0.081</u>	Test 6	<u>0.081</u>	Test 9	<u>0.081</u>
Average	<u>0.081</u>				

C. Remarks/Maintenance: ANNUAL INSPECTION

Instrument is acceptable to be used in the field.

Charles E. Edr  
Breath Analyst Signature

5.7.2019  
Date

NA

NA

Reviewed by

Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006500  
Location = HARV                  8164.13.00 06/09  
05/07/2019                                08:23

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-|=|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-|=|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&\*()\_+?

Current Instrument Setup

Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?         1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            34418080A2  
Standard Cyl #?            61  
Standard Expiration?       02/05/2021  
Oper No?                    666666  
  
Flow Cal. Date:             08/18/2015  
Slope                       681  
Intercept                  -680985

IR Calibration Date:        08/18/2015  
    3um                                9um

0th Coef(*100):	-3148	-17259
1st Coef(*100):	254475	138520
2nd Coef(*100):	1954	1329
H2O adj(mg/l*10k):	193	414

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
CHARLES EDER

Remarks:

PRINT TEST

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006500  
Location = HARV      8164.13.00 06/09  
05/07/2019      08:30

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:30
02 Std. Sol.	0.018	08:31
03 Room Air	0.000	08:32
04 Std. Sol.	0.018	08:32
05 Room Air	0.000	08:33
06 Std. Sol.	0.019	08:34
07 Room Air	0.000	08:34

08 Sim Temp = 34.0°C

Simul Ser No = DR3375

Std Sol No = 17160

County = 52

Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

Low AC  
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006500  
Location = HARV      8164.13.00 06/09  
05/07/2019      08:52

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:53
02 Std. Sol.	0.293	08:53
03 Room Air	0.000	08:54
04 Std. Sol.	0.296	08:55
05 Room Air	0.000	08:55
06 Std. Sol.	0.298	08:56
07 Room Air	0.000	08:57

08 Sim Temp = 34.0°C

Simul Ser No = DR7111

Std Sol No = 17350

County = 52

Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

HIGH AC  
0.300AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006500  
Location = HARV      8164.13.00 06/09  
05/07/2019      08:58

Test	AC	Time
01 Room Air	0.000	08:58
02 *Subject Test	INT*	08:59
03 Room Air	0.000	09:00

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = INTERFERENT  
Dr. Lic. = ND/DIS821456  
Lot No = 34418080A2  
Cyl No = 61  
Expiration Date = 02/05/2021  
County = 52      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*Interferent Check*  
*0.10 AC + 0.05% ACETONE*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006500  
Location = HARV                    8164.13.00 06/09  
05/07/2019                            09:00

Test	AC	Time
01 Diagnostic	OK	09:01
02 Room Air	0.000	09:02
03 *Subject Test	RFI*	09:02
04 Room Air	0.000	09:02

\*Invalid Test  
  Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = RFI CHECK  
Dr. Lic. = ND/DIS821456  
Lot No = 34418080A2  
Cyl No = 61  
Expiration Date = 02/05/2021  
County = 52                            Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks:

*RFI CHECK*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006500  
Location = HARV      8164.13.00 06/09  
05/07/2019      09:02

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:03
02 Std. Gas	0.080	09:03
03 Room Air	0.000	09:04
04 Std. Gas	0.081	09:04
05 Room Air	0.000	09:05
06 Std. Gas	0.081	09:05
07 Room Air	0.000	09:06

Lot No = 34418080A2  
Cyl No = 61  
Exp Date = 02/05/2021  
County = 52      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: **CALIBRATION CHECK**  
**0.080 AC**

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006500  
Location = HARV      8164.13.00 06/09  
05/07/2019      09:06

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:06
02 Std. Gas	0.081	09:07
03 Room Air	0.000	09:07
04 Std. Gas	0.080	09:08
05 Room Air	0.000	09:08
06 Std. Gas	0.081	09:09
07 Room Air	0.000	09:09

Lot No = 34418080A2  
Cyl No = 61  
Exp Date = 02/05/2021  
County = 52

Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

CALIBRATION CHECK  
0.080 AC

Form 106-I8000



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006500  
Location = HARV      8164.13.00 06/09  
05/07/2019      09:09

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:10
02 Std. Gas	0.081	09:10
03 Room Air	0.000	09:11
04 Std. Gas	0.081	09:11
05 Room Air	0.000	09:12
06 Std. Gas	0.081	09:12
07 Room Air	0.000	09:13

Lot No = 34418080A2  
Cyl No = 61  
Exp Date = 02/05/2021  
County = 52      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080 AC

Form 106-I8000