


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005946
Location = WILL 8164.14.00 09/16
05/08/2019 16:50

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:51
02 Std. Gas	0.081	16:51
03 Room Air	0.000	16:51
04 Std. Gas	0.080	16:52
05 Room Air	0.000	16:52
06 Std. Gas	0.079	16:53
07 Room Air	0.000	16:53

Lot No = 13518080A6
Cyl No = 28
Exp Date = 08/05/2020
County = 53 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK

0.080AC

Form 106-I8000