

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-005957

Location: DIVIDE COUNTY  
S.O. (CROS)

A. General Setup and Checks:

1.  Diagnostic Tests Pass and Instrument in Ready Mode
2.  Breath tube heated
3.  Date, time and location code (Level 2,E). Re-set if necessary.
4.  Print test (Level 1,P). Sign and attach test record.
5.  Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
  - a. Display 1004 psi Regulator: 1000 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a.  Low AC (Level 1,C): Known Value  $\leq 0.03$  AC: 0.020 AC  
Sim. Ser #: DR3375 Lot #: 17160 Exp. Date: 6.8.19
- b.  High AC (Level 1,C): Known Value  $\geq 0.25$  AC: 0.300 AC  
Sim. Ser #: DR7111 Lot #: 17350 Exp. Date: 10.11.19

Configure dry gas standard for the following tests (Level 1,S).

2.  Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone  
Sim. Ser #: DR7089 Lot #: ICS 5 Exp. Date: NA

Display reads "Interferent Detect"

3.  RFI Check (CMS Mode)  Display reads "RFI Detect"

4.  Dry Calibration Check (Level 1,C): Known Value 0.080 AC  
Gas Cyl Lot #: 34917080A3 Cyl #: 32 Exp. Date: 2.5.20

Test 1	<u>0.079</u>	Test 4	<u>0.079</u>	Test 7	<u>0.079</u>
Test 2	<u>0.080</u>	Test 5	<u>0.079</u>	Test 8	<u>0.079</u>
Test 3	<u>0.078</u>	Test 6	<u>0.080</u>	Test 9	<u>0.079</u>
Average <u>0.079</u>					

C. Remarks/Maintenance: ANNUAL INSPECTION

Instrument is acceptable to be used in the field.

Charles E. Eden  
Breath Analyst Signature

NA

5.8.2019  
Date

NA

Reviewed by

Date



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005957  
Location = CROS      8164.14.00 09/16  
05/08/2019      14:22

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:22
02 Std. Sol.	0.019	14:23
03 Room Air	0.000	14:24
04 Std. Sol.	0.019	14:24
05 Room Air	0.000	14:25
06 Std. Sol.	0.019	14:26
07 Room Air	0.000	14:26

08 Sim Temp = 34.0°C

Simul Ser No = DR3375

Std Sol No = 17160

County = 12

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

*Low AC  
0.020 AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005957  
Location = CROS      8164.14.00 09/16  
05/08/2019      14:28

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:28
02 Std. Sol.	0.297	14:29
03 Room Air	0.000	14:30
04 Std. Sol.	0.300	14:30
05 Room Air	0.000	14:31
06 Std. Sol.	0.301	14:32
07 Room Air	0.000	14:32

08 Sim Temp = 34.0°C

Simul Ser No = DR7111  
Std Sol No = 17350  
County = 12      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*HIGH AC*  
*0.300 AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005957  
Location = CROS                    8164.14.00 09/16  
05/08/2019                            14:12

Test	AC	Time
01 Room Air	0.000	14:13
02 *Subject Test	INT*	14:13
03 Room Air	0.000	14:14

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = INTERFERENT  
Dr. Lic. = ND/DIS821456  
Lot No = 34917080A3  
Cyl No = 32  
Expiration Date = 02/05/2020  
County = 12                            Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: *Interferent Check*  
*0.10 AC + 0.05% ACETONE*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005957  
Location = CROS      8164.14.00 09/16  
05/08/2019      14:18

Test	AC	Time
01 Diagnostic	OK	14:19
02 Room Air	0.000	14:20
03 *Subject Test	RFI*	14:20
04 Room Air	0.000	14:20

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = RFI CHECK  
Dr. Lic. = ND/DIS821456  
Lot No = 34917080A3  
Cyl No = 32  
Expiration Date = 02/05/2020  
County = 12      Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks:

*RFI CHECK*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005957  
Location = CROS      8164.14.00 09/16  
05/08/2019      14:08

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:09
02 Std. Gas	0.079	14:09
03 Room Air	0.000	14:10
04 Std. Gas	0.080	14:10
05 Room Air	0.000	14:10
06 Std. Gas	0.078	14:11
07 Room Air	0.000	14:11

Lot No = 34917080A3  
Cyl No = 32  
Exp Date = 02/05/2020  
County = 12      Oper No. = 666666

  
\_\_\_\_\_  
Operator Signature  
CHARLES EDER

Remarks: *CALIBRATION CHECK*  
*0.080 AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005957  
Location = CROS      8164.14.00 09/16  
05/08/2019      14:15

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:15
02 Std. Gas	0.079	14:16
03 Room Air	0.000	14:16
04 Std. Gas	0.079	14:17
05 Room Air	0.000	14:17
06 Std. Gas	0.080	14:18
07 Room Air	0.000	14:18

Lot No = 34917080A3  
Cyl No = 32  
Exp Date = 02/05/2020  
County = 12      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*CALIBRATION CHECK*

*0.080 AC*

Form 106-I8000



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005957  
Location = CROS      8164.14.00 09/16  
05/08/2019      14:33

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:34
02 Std. Gas	0.079	14:34
03 Room Air	0.000	14:35
04 Std. Gas	0.079	14:35
05 Room Air	0.000	14:36
06 Std. Gas	0.079	14:36
07 Room Air	0.000	14:37

Lot No = 34917080A3  
Cyl No = 32  
Exp Date = 02/05/2020  
County = 12      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080 AC

Form 106-I8000