

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-005937

Location: PEMBINA COUNTY  
S.O. (CAVA)

A. General Setup and Checks:

1.  Diagnostic Tests Pass and Instrument in Ready Mode
2.  Breath tube heated
3.  Date, time and location code (Level 2,E). Re-set if necessary.
4.  Print test (Level 1,P). Sign and attach test record.
5.  Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
  - a. Display: 86.8 psi Regulator: 875 psi

B. Tests (Sign and attach test records):

- Configure simulator for the following tests (Level 1,S).
1. Wet Calibration Check
    - a.  Low AC (Level 1,C): Known Value  $\leq 0.03$  AC: 0.020 AC  
Sim. Ser #: DR3375 Lot #: 17160 Exp. Date: 6.8.19
    - b.  High AC (Level 1,C): Known Value  $\geq 0.25$  AC: 0.300 AC  
Sim. Ser #: DR7111 Lot #: 17350 Exp. Date: 10.11.19

Configure dry gas standard for the following tests (Level 1,S).

2.  Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone  
Sim. Ser #: DR7089 Lot #: ICS 5 Exp. Date: NA  
 Display reads "Interferent Detect"
3.  RFI Check (CMS Mode)  Display reads "RFI Detect"
4.  Dry Calibration Check (Level 1,C): Known Value 0.080 AC  
Gas Cyl Lot #: 19817080A4 Cyl #: 3 Exp. Date: 9.5.19

Test 1	<u>0.079</u>	Test 4	<u>0.079</u>	Test 7	<u>0.079</u>
Test 2	<u>0.079</u>	Test 5	<u>0.080</u>	Test 8	<u>0.079</u>
Test 3	<u>0.080</u>	Test 6	<u>0.080</u>	Test 9	<u>0.079</u>
Average	<u>0.079</u>				

C. Remarks/Maintenance: REPLACED SIMULATOR RETURN O-RING,  
CEE ~~ANNUAL~~ ANNUAL INSPECTION  
REPLACED BREATH HOSE O-RING

Instrument is acceptable to be used in the field.

Charles E. CEE  
Breath Analyst Signature

5.13.19  
Date

NA  
Reviewed by

NA  
Date



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005937  
Location = CAVA      8164.14.00 09/16  
05/13/2019      15:43

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:43
02 Std. Sol.	0.019	15:44
03 Room Air	0.000	15:44
04 Std. Sol.	0.019	15:45
05 Room Air	0.000	15:46
06 Std. Sol.	0.019	15:46
07 Room Air	0.000	15:47


08 Sim Temp = 34.0°C

Simul Ser No = DR3375

Std Sol No = 17160

County = 34

Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*Low AC*  
*0.020 AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005937  
Location = CAVA      8164.14.00 09/16  
05/13/2019      15:48

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:48
02 Std. Sol.	0.301	15:49
03 Room Air	0.000	15:50
04 Std. Sol.	0.302	15:50
05 Room Air	0.000	15:51
06 Std. Sol.	0.303	15:52
07 Room Air	0.000	15:52

08 Sim Temp = 34.0°C

Simul Ser No = DR7111

Std Sol No = 17350

County = 34

Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

HIGH AC  
0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005937  
Location = CAVA      8164.14.00 09/16  
05/13/2019      15:26

Test	AC	Time
01 Room Air	0.000	15:27
02 *Subject Test	INT*	15:28
03 Room Air	0.000	15:28

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = INTERFERENT  
Dr. Lic. = ND/DIS821456  
Lot No = 19817080A4  
Cyl No = 3  
Expiration Date = 09/05/2019  
County = 34      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: *Interferent Check*  
*0.10 AC + 0.05% ACETONE*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005937  
Location = CAVA                    8164.14.00 09/16  
05/13/2019                            15:24

Test	AC	Time
01 Diagnostic	OK	15:25
02 Room Air	0.000	15:25
03 *Subject Test	RFI*	15:26
04 Room Air	0.000	15:26

\*Invalid Test  
  Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = RFI CHECK  
Dr. Lic. = ND/DIS821456  
Lot No = 19817080A4  
Cyl No = 3  
Expiration Date = 09/05/2019  
County = 34                            Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks:

*RFI CHECK*

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005937  
Location = CAVA      8164.14.00 09/16  
05/13/2019      15:20

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:21
02 Std. Gas	0.079	15:21
03 Room Air	0.000	15:21
04 Std. Gas	0.079	15:22
05 Room Air	0.000	15:22
06 Std. Gas	0.080	15:23
07 Room Air	0.000	15:23

Lot No = 19817080A4  
Cyl No = 3  
Exp Date = 09/05/2019  
County = 34      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005937  
Location = CAVA      8164.14.00 09/16  
05/13/2019      15:30

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:30
02 Std. Gas	0.079	15:30
03 Room Air	0.000	15:31
04 Std. Gas	0.080	15:31
05 Room Air	0.000	15:32
06 Std. Gas	0.080	15:32
07 Room Air	0.000	15:33

Lot No = 19817080A4  
Cyl No = 3  
Exp Date = 09/05/2019  
County = 34      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: *CALIBRATION CHECK*  
*0.080 AC*

Form 106-I8000




Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005937  
Location = CAVA      8164.14.00 09/16  
05/13/2019      15:54

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:55
02 Std. Gas	0.079	15:55
03 Room Air	0.000	15:56
04 Std. Gas	0.079	15:56
05 Room Air	0.000	15:57
06 Std. Gas	0.079	15:57
07 Room Air	0.000	15:57

Lot No = 19817080A4  
Cyl No = 3  
Exp Date = 09/05/2019  
County = 34      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080 AC

Form 106-I8000