



ETHANOL GAS STANDARD CYLINDER REPORT

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>CHARLES E. EDER, STATE TOX., ANNUAL INSPECTION</i>		
Location <i>GRAND FORKS AIR FORCE BASE (GFAF)</i>	Intoxilyzer® Serial Number <i>80-006497</i>	
Gas Lot Number <i>34418080A2</i>	Gas Cylinder Number <i>51</i>	Gas Expiration Date <i>2.5.2021</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. 0.*083* AC
 - B. 0.*082* AC
 - C. 0.*082* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>Charles E. Eder</i>	Date <i>5.15.19</i>
Reviewed By (Crime Laboratory Use Only) <i>Charles Eder</i>	Date <i>5.15.19</i>


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006497
Location = GFAF 8164.14.00 09/16
05/15/2019 08:22

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:22
02 Std. Gas	0.083	08:23
03 Room Air	0.000	08:23
04 Std. Gas	0.082	08:24
05 Room Air	0.000	08:24
06 Std. Gas	0.082	08:25
07 Room Air	0.000	08:25

Lot No = 34418080A2
Cyl No = 51
Exp Date = 02/05/2021
County = 65 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: **GAS STANDARD
CYLINDER INSTALL**

Form 106-I8000