



ETHANOL GAS STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (10-2013)

Chemical Test Operator Name (Print) <i>Danielle Bjork</i>		
Location <i>NCSD</i>	Intoxilyzer® Serial Number <i>80-006503</i>	
Gas Lot Number <i>34H18080AZ</i>	Gas Cylinder Number <i>47</i>	Gas Expiration Date <i>2-5-21</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C). *(ACA TEST PROVIDED)*
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 A. *0.081* AC
 B. *0.081* AC
 C. *0.080* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory: *(RECEIVED)*
 A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 B. ACA Test Record.

*CEE
6.10.19*

*CEE
6.10.19*

Chemical Test Operator Signature <i>[Signature]</i>	Date <i>6/4/19</i>
Reviewed By (Crime Laboratory Use Only) <i>[Signature]</i>	Date <i>6.10.2019</i>


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006503
Location = NCSO 8164.14.00 09/16
06/04/2019 08:07

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:07
02 Std. Gas	0.081	08:08
03 Room Air	0.000	08:08
04 Std. Gas	0.081	08:09
05 Room Air	0.000	08:09
06 Std. Gas	0.080	08:09
07 Room Air	0.000	08:10

Lot No = 34418080A2
Cyl No = 47
Exp Date = 02/05/2021
County = 32 Oper No. = 100754


Operator Signature
DANIELLE BJORLIE

Remarks:

Form 106-I8000