

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-005360

Location: TOXL

A. General Setup and Checks:

- 1. Diagnostic Tests Pass and Instrument in Ready Mode
- 2. Breath tube heated
- 3. Date, time and location code (Level 2,E). Re-set if necessary.
- 4. Print test (Level 1,P). Sign and attach test record.
- 5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 453 psi Regulator: 475 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a. Low AC (Level 1,C): Known Value \leq 0.03 AC: 0.026 AC
Sim. Ser #: DR3378 Lot #: 20172B Exp. Date: 12-12-19
- b. High AC (Level 1,C): Known Value \geq 0.25 AC: 0.300 AC
Sim. Ser #: DR7350 Lot #: 2018034 Exp. Date: 3-22-20

Configure dry gas standard for the following tests (Level 1,S).

- 2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR3451 Lot #: FCS5 Exp. Date: NA
 Display reads "Interferent Detect"
- 3. RFI Check (CMS Mode) Display reads "RFI Detect"
- 4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 34418086A2 Cyl #: 45 Exp. Date: 2-5-21

Test 1	<u>0.079</u>	Test 4	<u>0.080</u>	Test 7	<u>0.079</u>
Test 2	<u>0.079</u>	Test 5	<u>0.080</u>	Test 8	<u>0.080</u>
Test 3	<u>0.079</u>	Test 6	<u>0.079</u>	Test 9	<u>0.078</u>
Average <u>0.079</u>					

C. Remarks/Maintenance: ANNUAL INSPECTION

Instrument is acceptable to be used in the field.

Charles E. Edr
Breath Analyst Signature

6-11-19
Date

NA
Reviewed by

NA
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005360
Location = TOXL 8164.14.00 09/16
06/11/2019 13:47

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890- = |
ABCDEFGHIJKLMN OPQRSTUVWXYZ!@# \$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890- = |
ABCDEFGHIJKLMN OPQRSTUVWXYZ!@# \$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 34418080A2
Standard Cyl #? 45
Standard Expiration? 02/05/2021
Oper No? 666666

Flow Cal. Date: 08/17/2011
Slope 639
Intercept -343114

IR Calibration Date: 05/18/2017
 3um 9um

0th Coef(*100): -16956 -22484
1st Coef(*100): 266917 141716
2nd Coef(*100): 2089 1446
H2O adj(mg/l*10k): 216 365

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005360
Location = TOXL 8164.14.00 09/16
06/11/2019 13:50

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:51
02 Std. Sol.	0.018	13:52
03 Room Air	0.000	13:52
04 Std. Sol.	0.019	13:53
05 Room Air	0.000	13:54
06 Std. Sol.	0.019	13:54
07 Room Air	0.000	13:55

08 Sim Temp = 34.0°C

Simul Ser No = DR3378
Std Sol No = 201712B
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Low AC*
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005360
Location = TOXL 8164.14.00 09/16
06/11/2019 13:57

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:58
02 Std. Sol.	0.296	13:58
03 Room Air	0.000	13:59
04 Std. Sol.	0.299	14:00
05 Room Air	0.000	14:00
06 Std. Sol.	0.297	14:01
07 Room Air	0.000	14:01

08 Sim Temp = 34.0°C

Simul Ser No = DR7350

Std Sol No = 201803H

County = 08

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

High AC
0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005360
Location = TOXL 8164.14.00 09/16
06/11/2019 14:03

Test	AC	Time
01 Room Air	0.000	14:03
02 *Subject Test	INT*	14:04
03 Room Air	0.000	14:05

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS921456
Lot No = 34418080A2
Cyl No = 45
Expiration Date = 02/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent Check
0.10AC + 0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005360
Location = TOXL 8164.14.00 09/16
06/11/2019 14:05

Test	AC	Time
01 Diagnostic	OK	14:05
02 Room Air	0.000	14:06
03 *Subject Test	RFI*	14:06
04 Room Air	0.000	14:07

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS921456
Lot No = 34418080A2
Cyl No = 45
Expiration Date = 02/05/2021
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005360
Location = TOXL 8164.14.00 09/16
06/11/2019 14:07

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:07
02 Std. Gas	0.079	14:08
03 Room Air	0.000	14:08
04 Std. Gas	0.079	14:09
05 Room Air	0.000	14:09
06 Std. Gas	0.079	14:09
07 Room Air	0.000	14:10

Lot No = 34418080A2
Cyl No = 45
Exp Date = 02/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *CALIBRATION CHECK*
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005360
Location = TOXL 8164.14.00 09/16
06/11/2019 14:10

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:11
02 Std. Gas	0.080	14:11
03 Room Air	0.000	14:12
04 Std. Gas	0.080	14:12
05 Room Air	0.000	14:13
06 Std. Gas	0.079	14:13
07 Room Air	0.000	14:13

Lot No = 34418080A2
Cyl No = 45
Exp Date = 02/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005360
Location = TOXL 8164.14.00 09/16
06/11/2019 14:14

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:15
02 Std. Gas	0.079	14:15
03 Room Air	0.000	14:15
04 Std. Gas	0.080	14:16
05 Room Air	0.000	14:16
06 Std. Gas	0.078	14:17
07 Room Air	0.000	14:17

Lot No = 34418080A2
Cyl No = 45
Exp Date = 02/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000