

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-005943

Location: TOXL

A. General Setup and Checks:

1.  Diagnostic Tests Pass and Instrument in Ready Mode
2.  Breath tube heated
3.  Date, time and location code (Level 2,E). Re-set if necessary.
4.  Print test (Level 1,P). Sign and attach test record.
5.  Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
  - a. Display: 473 psi Regulator: 500 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a.  Low AC (Level 1,C): Known Value  $\leq 0.03$  AC: 0.020 AC  
Sim. Ser #: DR3378 Lot #: 201712B Exp. Date: 12-12-19
- b.  High AC (Level 1,C): Known Value  $\geq 0.25$  AC: 0.300 AC  
Sim. Ser #: DR7350 Lot #: 201803H Exp. Date: 3-22-20

Configure dry gas standard for the following tests (Level 1,S).

2.  Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone  
Sim. Ser #: DR3451 Lot #: ICS 5 Exp. Date: NA  
 Display reads "Interferent Detect"
3.  RFI Check (CMS Mode)  Display reads "RFI Detect"
4.  Dry Calibration Check (Level 1,C): Known Value 0.080 AC  
Gas Cyl Lot #: 344180B0A2 Cyl #: 45 Exp. Date: 2-5-21

Test 1	<u>0.079</u>	Test 4	<u>0.080</u>	Test 7	<u>0.080</u>
Test 2	<u>0.080</u>	Test 5	<u>0.079</u>	Test 8	<u>0.080</u>
Test 3	<u>0.080</u>	Test 6	<u>0.080</u>	Test 9	<u>0.080</u>
Average	<u>0.080</u>		<u>0.080</u>		

C. Remarks/Maintenance: ANNUAL INSPECTION

Instrument is acceptable to be used in the field.

Charles E. Ed  
Breath Analyst Signature

6-11-2019  
Date

NA  
Reviewed by

NA  
Date



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005943  
Location = TOXL      8164.14.00 09/16  
06/11/2019      13:11

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:11
02 Std. Sol.	0.018	13:12
03 Room Air	0.000	13:13
04 Std. Sol.	0.019	13:13
05 Room Air	0.000	13:14
06 Std. Sol.	0.019	13:15
07 Room Air	0.000	13:15

08 Sim Temp = 34.0°C

Simul Ser No = DR3378  
Std Sol No = 201712B  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*Low AC*  
*0.020 AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005943  
Location = TOXL      8164.14.00 09/16  
06/11/2019      13:16

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:17
02 Std. Sol.	0.297	13:18
03 Room Air	0.000	13:18
04 Std. Sol.	0.299	13:19
05 Room Air	0.000	13:19
06 Std. Sol.	0.299	13:20
07 Room Air	0.000	13:21

08 Sim Temp = 34.0°C

Simul Ser No = DR7350  
Std Sol No = 201803H  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: HIGH AC  
0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005943  
Location = TOXL                 8164.14.00 09/16  
06/11/2019                        13:22

Test	AC	Time
01 Room Air	0.000	13:23
02 *Subject Test	INT*	13:23
03 Room Air	0.000	13:24

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                 Weight = 150  
Test = DUI                         Cit = INTERFERENT  
Dr. Lic. = ND/DIS921456  
Lot No = 34418080A2  
Cyl No = 45  
Expiration Date = 02/05/2021  
County = 08                        Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: *Interferent Check*  
*0.10 AC + 0.05% ACETONE*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005943  
Location = TOXL      8164.14.00 09/16  
06/11/2019      13:24

Test	AC	Time
01 Diagnostic	OK	13:25
02 Room Air	0.000	13:25
03 *Subject Test	RFI*	13:25
04 Room Air	0.000	13:26

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = RFI CHECK  
Dr. Lic. = ND/DIS921456  
Lot No = 34418080A2  
Cyl No = 45  
Expiration Date = 02/05/2021  
County = 08      Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks:

*RFI CHECK*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005943  
Location = TOXL                    8164.14.00 09/16  
06/11/2019                            13:26

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:27
02 Std. Gas	0.079	13:27
03 Room Air	0.000	13:28
04 Std. Gas	0.080	13:28
05 Room Air	0.000	13:29
06 Std. Gas	0.080	13:29
07 Room Air	0.000	13:30

Lot No = 34418080A2  
Cyl No = 45  
Exp Date = 02/05/2021  
County = 08                            Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005943  
Location = TOXL      8164.14.00 09/16  
06/11/2019      13:30

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:30
02 Std. Gas	0.080	13:31
03 Room Air	0.000	13:31
04 Std. Gas	0.079	13:32
05 Room Air	0.000	13:32
06 Std. Gas	0.080	13:33
07 Room Air	0.000	13:33

Lot No = 34418080A2  
Cyl No = 45  
Exp Date = 02/05/2021  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005943  
Location = TOXL      8164.14.00 09/16  
06/11/2019      13:33

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:34
02 Std. Gas	0.080	13:34
03 Room Air	0.000	13:35
04 Std. Gas	0.080	13:35
05 Room Air	0.000	13:36
06 Std. Gas	0.080	13:36
07 Room Air	0.000	13:37

Lot No = 34418080A2  
Cyl No = 45  
Exp Date = 02/05/2021  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080 AC

Form 106-I8000