

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-004936

Location: TOXL

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 422 psi Regulator: 450 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: DR337B Lot #: 201712B Exp. Date: 12.12.19
- b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.300 AC
Sim. Ser #: DR7350 Lot #: 2018034 Exp. Date: 3.22.20

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR3451 Lot #: ICS 5 Exp. Date: NA

Display reads "Interferent Detect"

3. RFI Check (CMS Mode) Display reads "RFI Detect"

4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC

Gas Cyl Lot #: 34418080A2 Cyl #: 45 Exp. Date: 2.5.21

Test 1 0.080 Test 4 0.079 Test 7 0.080

Test 2 0.079 Test 5 0.080 Test 8 0.080

Test 3 0.080 Test 6 0.080 Test 9 0.079

Average 0.080

C. Remarks/Maintenance: ANNUAL INSPECTION

Instrument is acceptable to be used in the field.

Charles E. Ehr
Breath Analyst Signature

6/14/19
Date

NA
Reviewed by

NA
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004936
Location = TOXL 8164.14.00 09/16
06/14/2019 16:38

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 34418080A2
Standard Cyl #? 45
Standard Expiration? 02/05/2021
Oper No? 666666

Flow Cal. Date: 01/18/2010
Slope 732
Intercept -767769

IR Calibration Date: 04/20/2016
 3um 9um

0th Coef(*100): -15846 -19883
1st Coef(*100): 260423 131703
2nd Coef(*100): 1755 1110
H2O adj(mg/l*10k): 371 394

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004936
Location = TOXL 8164.14.00 09/16
06/14/2019 16:39

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:40
02 Std. Sol.	0.019	16:41
03 Room Air	0.000	16:41
04 Std. Sol.	0.019	16:42
05 Room Air	0.000	16:43
06 Std. Sol.	0.019	16:43
07 Room Air	0.000	16:44

08 Sim Temp = 34.0°C

Simul Ser No = DR3378
Std Sol No = 201712B
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Low AC*
 0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004936
Location = TOXL 8164.14.00 09/16
06/14/2019 16:49

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:50
02 Std. Sol.	0.297	16:51
03 Room Air	0.000	16:51
04 Std. Sol.	0.296	16:52
05 Room Air	0.000	16:52
06 Std. Sol.	0.296	16:53
07 Room Air	0.000	16:54

08 Sim Temp = 34.0°C

Simul Ser No = DR7350
Std Sol No = 201803H
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

HIGH AC
0.300 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004936
Location = TOXL 8164.14.00 09/16
06/14/2019 16:34

Test	AC	Time
01 Room Air	0.000	16:35
02 *Subject Test	INT*	16:35
03 Room Air	0.000	16:36

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS921456
Lot No = 34418080A2
Cyl No = 45
Expiration Date = 02/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent Check
0.10 AC + 0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004936
Location = TOXL 8164.14.00 09/16
06/14/2019 16:36

Test	AC	Time
01 Diagnostic	OK	16:37
02 Room Air	0.000	16:37
03 *Subject Test	RFI*	16:37
04 Room Air	0.000	16:38

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS921456
Lot No = 34418080A2
Cyl No = 45
Expiration Date = 02/05/2021
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004936
Location = TOXL 8164.14.00 09/16
06/14/2019 16:55

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:55
02 Std. Gas	0.080	16:56
03 Room Air	0.000	16:56
04 Std. Gas	0.079	16:56
05 Room Air	0.000	16:57
06 Std. Gas	0.080	16:57
07 Room Air	0.000	16:58

Lot No = 34418080A2
Cyl No = 45
Exp Date = 02/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004936
Location = TOXL 8164.14.00 09/16
06/14/2019 16:58

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:59
02 Std. Gas	0.079	16:59
03 Room Air	0.000	16:59
04 Std. Gas	0.080	17:00
05 Room Air	0.000	17:00
06 Std. Gas	0.080	17:01
07 Room Air	0.000	17:01

Lot No = 34418080A2
Cyl No = 45
Exp Date = 02/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004936
Location = TOXL 8164.14.00 09/16
06/14/2019 17:01

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:02
02 Std. Gas	0.080	17:02
03 Room Air	0.000	17:03
04 Std. Gas	0.080	17:03
05 Room Air	0.000	17:04
06 Std. Gas	0.079	17:04
07 Room Air	0.000	17:05

Lot No = 34418080A2
Cyl No = 45
Exp Date = 02/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000