

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-006501

Location: TOXL

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 197 psi Regulator: 225 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check
 - a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: DR3378 Lot #: 201712B Exp. Date: 12-12-19
 - b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.250 AC
Sim. Ser #: DR7351 Lot #: 201803G Exp. Date: 3-22-20

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR3451 Lot #: ICS 5 Exp. Date: NA
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 34418080A2 Cyl #: 45 Exp. Date: 2-25-21

Test 1 <u>0.080</u>	Test 4 <u>0.080</u>	Test 7 <u>0.080</u> CEE
Test 2 <u>0.080</u>	Test 5 <u>0.080</u>	Test 8 <u>0.080</u>
Test 3 <u>0.080</u>	Test 6 <u>0.080</u>	Test 9 <u>0.080</u>
Average <u>0.080</u>		

C. Remarks/Maintenance: ANNUAL INSPECTION

Instrument is acceptable to be used in the field.

Charles E. Eel
Breath Analyst Signature

6/14/2019
Date

NA

NA

Reviewed by _____

Date _____

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006501
Location = TOXL 8164.14.00 09/16
06/14/2019 12:50

***** Printer Test *****

abcdefghijklmnopqrstu vwxyz1234567890-=-|
ABCDEFGHIJKLMN O PQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstu vwxyz1234567890-=-|
ABCDEFGHIJKLMN O PQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup
Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 34418080A2
Standard Cyl #? 45
Standard Expiration? 02/05/2021
Oper No? 666666

Flow Cal. Date: 06/13/2019
Slope 651
Intercept -777821

IR Calibration Date: 06/13/2019

	3um	9um
0th Coef(*100):	-20531	-14906
1st Coef(*100):	270252	139749
2nd Coef(*100):	1917	1284
H2O adj(mg/l*10k):	393	321

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks: *PRINT TEST*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006501
Location = TOXL 8164.14.00 09/16
06/14/2019 12:51

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:52
02 Std. Sol.	0.020	12:53
03 Room Air	0.000	12:53
04 Std. Sol.	0.020	12:54
05 Room Air	0.000	12:54
06 Std. Sol.	0.020	12:55
07 Room Air	0.000	12:56

08 Sim Temp = 34.0°C

Simul Ser No = DR3378

Std Sol No = 201712B

County = 08

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Low AC
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006501
Location = TOXL 8164.14.00 09/16
06/14/2019 12:57

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:58
02 Std. Sol.	0.247	12:58
03 Room Air	0.000	12:59
04 Std. Sol.	0.249	13:00
05 Room Air	0.000	13:00
06 Std. Sol.	0.249	13:01
07 Room Air	0.000	13:02

08 Sim Temp = 34.0°C

Simul Ser No = DR7351
Std Sol No = 201803G
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

HIGH AC
0.250 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006501
Location = TOXL 8164.14.00 09/16
06/14/2019 13:03

Test	AC	Time
01 Room Air	0.000	13:04
02 *Subject Test	INT*	13:04
03 Room Air	0.000	13:05

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS921456
Lot No = 34418080A2
Cyl No = 45
Expiration Date = 02/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent Check
0.10AC + 0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006501
Location = TOXL 8164.14.00 09/16
06/14/2019 13:05

Test	AC	Time
01 Diagnostic	OK	13:06
02 Room Air	0.000	13:07
03 *Subject Test	RFI*	13:07
04 Room Air	0.000	13:07

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS921456
Lot No = 34418080A2
Cyl No = 45
Expiration Date = 02/05/2021
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: ~~OK~~ RFI CHECK
CEE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006501
Location = TOXL 8164.14.00 09/16
06/14/2019 13:07

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:08
02 Std. Gas	0.080	13:08
03 Room Air	0.000	13:09
04 Std. Gas	0.080	13:09
05 Room Air	0.000	13:10
06 Std. Gas	0.080	13:10
07 Room Air	0.000	13:11

Lot No = 34418080A2
Cyl No = 45
Exp Date = 02/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006501
Location = TOXL 8164.14.00 09/16
06/14/2019 13:11

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:12
02 Std. Gas	0.080	13:12
03 Room Air	0.000	13:13
04 Std. Gas	0.080	13:13
05 Room Air	0.000	13:13
06 Std. Gas	0.080	13:14
07 Room Air	0.000	13:14

Lot No = 34418080A2
Cyl No = 45
Exp Date = 02/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006501
Location = TOXL 8164.14.00 09/16
06/14/2019 13:15

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:15
02 Std. Gas	0.080	13:16
03 Room Air	0.000	13:16
04 Std. Gas	0.080	13:16
05 Room Air	0.000	13:17
06 Std. Gas	0.080	13:17
07 Room Air	0.000	13:18

Lot No = 34418080A2
Cyl No = 45
Exp Date = 02/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080AC

Form 106-I8000