

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-004946

Location: TOXL

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 491 psi Regulator: 500 psi

B. Tests (Sign and attach test records):

- Configure simulator for the following tests (Level 1,S).
1. Wet Calibration Check
 - a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: DR3378 Lot #: 201712B Exp. Date: 12.12.19
 - b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.300 AC
Sim. Ser #: DR7089 Lot #: 17350 Exp. Date: 10.11.19

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR3451 Lot #: ICS 5 Exp. Date: NA
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 34418080A2 Cyl #: 45 Exp. Date: 2.5.21

Test 1	<u>0.080</u>	Test 4	<u>0.080</u>	Test 7	<u>0.081</u>
Test 2	<u>0.081</u>	Test 5	<u>0.080</u>	Test 8	<u>0.080</u>
Test 3	<u>0.080</u>	Test 6	<u>0.079</u>	Test 9	<u>0.080</u>
Average	<u>0.080</u>				

C. Remarks/Maintenance: ANNUAL INSPECTION

Instrument is acceptable to be used in the field.

Charles E. Edr
Breath Analyst Signature

6.17.2019
Date

NA

NA

Reviewed by

Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004946
Location = TOXL 8164.14.00 09/16
06/17/2019 16:03

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:03
02 Std. Sol.	0.017	16:04
03 Room Air	0.000	16:05
04 Std. Sol.	0.018	16:05
05 Room Air	0.000	16:06
06 Std. Sol.	0.018	16:06
07 Room Air	0.000	16:07

08 Sim Temp = 34.0°C

Simul Ser No = DR3378
Std Sol No = 201712B
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Low AC
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004946
Location = TOXL 8164.14.00 09/16
06/17/2019 16:08

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:08
02 Std. Sol.	0.293	16:09
03 Room Air	0.000	16:10
04 Std. Sol.	0.295	16:10
05 Room Air	0.000	16:11
06 Std. Sol.	0.294	16:12
07 Room Air	0.000	16:12

08 Sim Temp = 34.0°C

Simul Ser No = DR7089
Std Sol No = 17350
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: HIGH AC
0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc., Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004946
Location = TOXL 8164.14.00 09/16
06/17/2019 16:19

Test	AC	Time
01 Room Air	0.000	16:20
02 *Subject Test	INT*	16:20
03 Room Air	0.000	16:21

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS921456
Lot No = 34418080A2
Cyl No = 45
Expiration Date = 02/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent Check

*0.10 AC + ~~0.05~~ 0.05% ACETONE
CEE*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004946
Location = TOXL 8164.14.00 09/16
06/17/2019 16:21

Test	AC	Time
01 Diagnostic	OK	16:22
02 Room Air	0.000	16:23
03 *Subject Test	RFI*	16:23
04 Room Air	0.000	16:23

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS921456
Lot No = 34418080A2
Cyl No = 45
Expiration Date = 02/05/2021
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004946
Location = TOXL 8164.14.00 09/16
06/17/2019 16:24

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:25
02 Std. Gas	0.080	16:25
03 Room Air	0.000	16:25
04 Std. Gas	0.081	16:26
05 Room Air	0.000	16:26
06 Std. Gas	0.080	16:27
07 Room Air	0.000	16:27

Lot No = 34418080A2
Cyl No = 45
Exp Date = 02/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004946
Location = TOXL 8164.14.00 09/16
06/17/2019 16:28

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:28
02 Std. Gas	0.080	16:29
03 Room Air	0.000	16:29
04 Std. Gas	0.080	16:29
05 Room Air	0.000	16:30
06 Std. Gas	0.079	16:30
07 Room Air	0.000	16:31

Lot No = 34418080A2
Cyl No = 45
Exp Date = 02/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004946
Location = TOXL 8164.14.00 09/16
06/17/2019 16:32

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:32
02 Std. Gas	0.081	16:33
03 Room Air	0.000	16:33
04 Std. Gas	0.080	16:34
05 Room Air	0.000	16:34
06 Std. Gas	0.080	16:34
07 Room Air	0.000	16:35

Lot No = 34418080A2
Cyl No = 45
Exp Date = 02/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000