

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-004943

Location: TOXL

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 460 psi Regulator: 450 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check
 - a. Low AC (Level 1,C): Known Value \leq 0.03 AC: 0.020 AC
Sim. Ser #: DR3378 Lot #: 201712B Exp. Date: 12.12.19
 - b. High AC (Level 1,C): Known Value \geq 0.25 AC: 0.300 AC
Sim. Ser #: DR7350 Lot #: 201803H Exp. Date: 3.22.20

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR3451 Lot #: ICS 5 Exp. Date: NA
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 34418080A2 Cyl #: 45 Exp. Date: 2.5.21

Test 1 <u>0.080</u>	Test 4 <u>0.081</u>	Test 7 <u>0.080</u>
Test 2 <u>0.081</u>	Test 5 <u>0.080</u>	Test 8 <u>0.081</u>
Test 3 <u>0.081</u>	Test 6 <u>0.081</u>	Test 9 <u>0.080</u>
Average <u>0.081</u>		

C. Remarks/Maintenance: ANNUAL INSPECTION

Instrument is acceptable to be used in the field.

Charles E. Edr
Breath Analyst Signature

6.17.2019
Date

NA
Reviewed by

NA
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004943
Location = TOXL 8164.14.00 09/16
06/17/2019 15:16

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:16
02 Std. Sol.	0.018	15:17
03 Room Air	0.000	15:18
04 Std. Sol.	0.018	15:18
05 Room Air	0.000	15:19
06 Std. Sol.	0.018	15:19
07 Room Air	0.000	15:20


08 Sim Temp = 34.0°C

Simul Ser No = DR3378

Std Sol No = 201712B

County = 08

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

Low AC

~~0.018~~ 0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004943
Location = TOXL 8164.14.00 09/16
06/17/2019 15:21

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:22
02 Std. Sol.	0.298	15:23
03 Room Air	0.000	15:23
04 Std. Sol.	0.299	15:24
05 Room Air	0.000	15:24
06 Std. Sol.	0.300	15:25
07 Room Air	0.000	15:26

08 Sim Temp = 34.0°C

Simul Ser No = DR7350
Std Sol No = 201803H
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: HIGH AC
0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004943
Location = TOXL 8164.14.00 09/16
06/17/2019 15:28

Test	AC	Time
01 Room Air	0.000	15:28
02 *Subject Test	INT*	15:29
03 Room Air	0.000	15:29

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS921456
Lot No = 34418080A2
Cyl No = 45
Expiration Date = 02/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Interferent Check*
0.10AC + 0.05% ACETONE
Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004943
Location = TOXL 8164.14.00 09/16
06/17/2019 15:32

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:33
02 Std. Gas	0.080	15:33
03 Room Air	0.000	15:34
04 Std. Gas	0.081	15:34
05 Room Air	0.000	15:34
06 Std. Gas	0.081	15:35
07 Room Air	0.000	15:35

Lot No = 34418080A2
Cyl No = 45
Exp Date = 02/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004943
Location = TOXL 8164.14.00 09/16
06/17/2019 15:36

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:36
02 Std. Gas	0.081	15:36
03 Room Air	0.000	15:37
04 Std. Gas	0.080	15:37
05 Room Air	0.000	15:38
06 Std. Gas	0.081	15:38
07 Room Air	0.000	15:39

Lot No = 34418080A2
Cyl No = 45
Exp Date = 02/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.090 AC
Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004943
Location = TOXL 8164.14.00 09/16
06/17/2019 15:39

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:40
02 Std. Gas	0.080	15:40
03 Room Air	0.000	15:41
04 Std. Gas	0.081	15:41
05 Room Air	0.000	15:41
06 Std. Gas	0.080	15:42
07 Room Air	0.000	15:42

Lot No = 34418080A2
Cyl No = 45
Exp Date = 02/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000