

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-004194

Location: TOX

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 425 psi Regulator: 450 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: DR3378 Lot #: 201712B Exp. Date: 12.12.19
- b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.300 AC
Sim. Ser #: DR7350 Lot #: 2018034 Exp. Date: 3.22.20

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR3451 Lot #: ICS 5 Exp. Date: NA
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 34418080 A2 Cyl #: 40 Exp. Date: 2.5.21

Test 1	<u>0.080</u>	Test 4	<u>0.079</u>	Test 7	<u>0.079</u>
Test 2	<u>0.080</u>	Test 5	<u>0.080</u>	Test 8	<u>0.079</u>
Test 3	<u>0.080</u>	Test 6	<u>0.079</u>	Test 9	<u>0.079</u>
Average	<u>0.079</u>				

C. Remarks/Maintenance: ANNUAL INSPECTION

Instrument is acceptable to be used in the field.

Charles E. Eden
Breath Analyst Signature

6.21.19
Date

NA

NA

Reviewed by

Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004194
Location = TOXL 8164.14.00 09/16
06/21/2019 14:33

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 34418080A2
Standard Cyl #? 40
Standard Expiration? 02/05/2021
Oper No? 666666

Flow Cal. Date: 06/23/2009
Slope 681
Intercept -622430

IR Calibration Date: 05/04/2016
 3um 9um

0th Coef(*100): -17808 -24119
1st Coef(*100): 263893 132021
2nd Coef(*100): 3080 1296
H2O adj(mg/l*10k): 403 519

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004194
Location = TOXL 8164.14.00 09/16
06/21/2019 14:33

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:34
02 Std. Sol.	0.017	14:35
03 Room Air	0.000	14:35
04 Std. Sol.	0.018	14:36
05 Room Air	0.000	14:36
06 Std. Sol.	0.017	14:37
07 Room Air	0.000	14:38

08 Sim Temp = 34.0°C

Simul Ser No = DR3378
Std Sol No = 201712B
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Low AC
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004194
Location = TOXL 8164.14.00 09/16
06/21/2019 14:38

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:39
02 Std. Sol.	0.295	14:40
03 Room Air	0.000	14:40
04 Std. Sol.	0.297	14:41
05 Room Air	0.000	14:41
06 Std. Sol.	0.297	14:42
07 Room Air	0.000	14:43


08 Sim Temp = 34.0°C

Simul Ser No = DR7350

Std Sol No = 201803H

County = 08

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

High AC
0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004194
Location = TOXL 8164.14.00 09/16
06/21/2019 14:44

Test	AC	Time
01 Room Air	0.000	14:45
02 *Subject Test	INT*	14:45
03 Room Air	0.000	14:46

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS921456
Lot No = 34418080A2
Cyl No = 40
Expiration Date = 02/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

~~RFI CHECK~~ CEE
Interferent Check

Form 106-I8000

0.10AC + 0.05% ACETONE

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004194
Location = TOXL 8164.14.00 09/16
06/21/2019 14:46

Test	AC	Time
01 Diagnostic	OK	14:47
02 Room Air	0.000	14:48
03 *Subject Test	RFI*	14:48
04 Room Air	0.000	14:48

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS921456
Lot No = 34418080A2
Cyl No = 40
Expiration Date = 02/05/2021
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004194
Location = TOXL 8164.14.00 09/16
06/21/2019 14:48

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:49
02 Std. Gas	0.080	14:49
03 Room Air	0.000	14:50
04 Std. Gas	0.080	14:50
05 Room Air	0.000	14:51
06 Std. Gas	0.080	14:51
07 Room Air	0.000	14:52

Lot No = 34418080A2
Cyl No = 40
Exp Date = 02/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *CALIBRATION CHECK*
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004194
Location = TOXL 8164.14.00 09/16
06/21/2019 14:52

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:52
02 Std. Gas	0.079	14:53
03 Room Air	0.000	14:53
04 Std. Gas	0.080	14:54
05 Room Air	0.000	14:54
06 Std. Gas	0.079	14:55
07 Room Air	0.000	14:55

Lot No = 34418080A2
Cyl No = 40
Exp Date = 02/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004194
Location = TOXL 8164.14.00 09/16
06/21/2019 14:55

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:56
02 Std. Gas	0.079	14:57
03 Room Air	0.000	14:57
04 Std. Gas	0.079	14:57
05 Room Air	0.000	14:58
06 Std. Gas	0.079	14:58
07 Room Air	0.000	14:59

Lot No = 34418080A2
Cyl No = 40
Exp Date = 02/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION CHECK
0.080 AC

Form 106-I8000