

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-003059

Location: TOXL

A. General Setup and Checks:

1.  Diagnostic Tests Pass and Instrument in Ready Mode
2.  Breath tube heated
3.  Date, time and location code (Level 2,E). Re-set if necessary.
4.  Print test (Level 1,P). Sign and attach test record.
5.  Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
  - a. Display: 436 psi Regulator: 450 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check
  - a.  Low AC (Level 1,C): Known Value  $\leq 0.03$  AC: 0.020 AC  
Sim. Ser #: DR3378 Lot #: 201712B Exp. Date: 12-12-19
  - b.  High AC (Level 1,C): Known Value  $\geq 0.25$  AC: 0.300 AC  
Sim. Ser #: DR7350 Lot #: 201803H Exp. Date: 3-22-20

Configure dry gas standard for the following tests (Level 1,S).

2.  Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone  
Sim. Ser #: DR3451 Lot #: ICS 5 Exp. Date: NA  
 Display reads "Interferent Detect"
3.  RFI Check (CMS Mode)  Display reads "RFI Detect"
4.  Dry Calibration Check (Level 1,C): Known Value 0.080 AC  
Gas Cyl Lot #: 3441808 0A2 Cyl #: 40 Exp. Date: 2-5-21

Test 1	<u>0.083</u>	Test 4	<u>0.081</u>	Test 7	<u>0.081</u>
Test 2	<u>0.082</u>	Test 5	<u>0.081</u>	Test 8	<u>0.081</u>
Test 3	<u>0.082</u>	Test 6	<u>0.082</u>	Test 9	<u>0.081</u>
Average <u>0.082</u>					

C. Remarks/Maintenance: ANNUAL INSPECTION

Instrument is acceptable to be used in the field.

Charles E. Ed  
Breath Analyst Signature

6.21.19  
Date

NA  
Reviewed by

NA  
Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-003059  
Location = TOXL                8164.14.00 09/16  
06/21/2019                      13:44

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890- = |  
ABCDEFGHIJKLMN OPQRSTUVWXYZ!@# \$%^&\* ( ) \_ + ?

abcdefghijklmnopqrstuvwxyz1234567890- = |  
ABCDEFGHIJKLMN OPQRSTUVWXYZ!@# \$%^&\* ( ) \_ + ?

Current Instrument Setup


Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?         1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            34418080A2  
Standard Cyl #?            40  
Standard Expiration?       02/05/2021  
Oper No?                    666666

Flow Cal. Date:            11/28/2018  
Slope                        589  
Intercept                  -276335

IR Calibration Date:       12/04/2018  
                                    3um                      9um

-----  
0th Coef(\*100):            -25330                    -15825  
1st Coef(\*100):            278286                    137135  
2nd Coef(\*100):            1992                      1257  
H2O adj(mg/l\*10k):        525                        321

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
CHARLES EDER

Remarks:

PRINT TEST

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003059  
Location = TOXL      8164.14.00 09/16  
06/21/2019      13:46

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:47
02 Std. Sol.	0.018	13:48
03 Room Air	0.000	13:48
04 Std. Sol.	0.018	13:49
05 Room Air	0.000	13:49
06 Std. Sol.	0.018	13:50
07 Room Air	0.000	13:51

08 Sim Temp = 34.0°C

Simul Ser No = DR3378  
Std Sol No = 201712B  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

Low AC  
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003059  
Location = TOXL      8164.14.00 09/16  
06/21/2019      13:56

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:57
02 Std. Sol.	0.296	13:58
03 Room Air	0.000	13:58
04 Std. Sol.	0.295	13:59
05 Room Air	0.000	13:59
06 Std. Sol.	0.296	14:00
07 Room Air	0.000	14:01

08 Sim Temp = 34.0°C

Simul Ser No = DR7350

Std Sol No = 201803H

County = 08

Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

HIGH AC  
0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-003059  
Location = TOXL                    8164.14.00 09/16  
06/21/2019                            14:02

Test	AC	Time
01 Room Air	0.000	14:03
02 *Subject Test	INT*	14:03
03 Room Air	0.000	14:04

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = INTERFERENT  
Dr. Lic. = ND/DIS921456  
Lot No = 34418080A2  
Cyl No = 40  
Expiration Date = 02/05/2021  
County = 08                            Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*Interferent Check*  
*0.10 AC + 0.05% ACETONE*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003059  
Location = TOXL      8164.14.00 09/16  
06/21/2019      14:04

Test	AC	Time
01 Diagnostic	OK	14:05
02 Room Air	0.000	14:05
03 Reference	RFI*	14:05
04 Room Air	0.000	14:06

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = RFI CHECK  
Dr. Lic. = ND/DIS921456  
Lot No = 34418080A2  
Cyl No = 40  
Expiration Date = 02/05/2021  
County = 08      Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks: *RFI CHECK*

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003059  
Location = TOXL      8164.14.00 09/16  
06/21/2019      14:06

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:07
02 Std. Gas	0.083	14:07
03 Room Air	0.000	14:08
04 Std. Gas	0.082	14:08
05 Room Air	0.000	14:08
06 Std. Gas	0.082	14:09
07 Room Air	0.000	14:09

Lot No = 34418080A2  
Cyl No = 40  
Exp Date = 02/05/2021  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003059  
Location = TOXL      8164.14.00 09/16  
06/21/2019      14:09

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:10
02 Std. Gas	0.081	14:10
03 Room Air	0.000	14:11
04 Std. Gas	0.081	14:11
05 Room Air	0.000	14:12
06 Std. Gas	0.082	14:12
07 Room Air	0.000	14:13

Lot No = 34418080A2  
Cyl No = 40  
Exp Date = 02/05/2021  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080 AC

Form 106-I8000



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003059  
Location = TOXL      8164.14.00 09/16  
06/21/2019      14:13

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:13
02 Std. Gas	0.081	14:14
03 Room Air	0.000	14:14
04 Std. Gas	0.081	14:15
05 Room Air	0.000	14:15
06 Std. Gas	0.081	14:15
07 Room Air	0.000	14:16

Lot No = 34418080A2  
Cyl No = 40  
Exp Date = 02/05/2021  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*CALIBRATION CHECK*  
*0.080AC*

Form 106-I8000