

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-005941

Location: TOXL

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 434 psi Regulator: 450 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check
 - a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: DR3378 Lot #: 18020 Exp. Date: 1.9.20
 - b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.300 AC
Sim. Ser #: DR7350 Lot #: 2018034 Exp. Date: 3.22.20

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR3451 Lot #: ICS 5 Exp. Date: NA
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 34418080 A2 Cyl #: 40 Exp. Date: 2.5.21

Test 1	<u>0.081</u>	Test 4	<u>0.081</u>	Test 7	<u>0.081</u>
Test 2	<u>0.082</u>	Test 5	<u>0.081</u>	Test 8	<u>0.081</u>
Test 3	<u>0.081</u>	Test 6	<u>0.081</u>	Test 9	<u>0.081</u>
Average	<u>CEE 0.081</u>				

C. Remarks/Maintenance: ANNUAL INSPECTION

NOTE: INSTRUMENT IS CLASSROOM ONLY, AUTO CAL RANGE IS LOW

Instrument is acceptable to be used in the field.

Charles E. Ed
Breath Analyst Signature

6.24.2019
Date

NA
Reviewed by

NA
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005941
Location = TOXL 8164.14.00 09/16
06/24/2019 14:44

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&*()_+?

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 34418080A2
Standard Cyl #? 40
Standard Expiration? 02/05/2021
Oper No? 666666

Flow Cal. Date: 05/30/2018
Slope 618
Intercept -570330

IR Calibration Date: 05/30/2018
 3um 9um

0th Coef(*100): -13785 -15807
1st Coef(*100): 281869 139821
2nd Coef(*100): 2733 1395
H2O adj(mg/l*10k): 308 362

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005941
Location = TOXL 8164.14.00 09/16
06/24/2019 14:48

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:48
02 Std. Sol.	0.019	14:49
03 Room Air	0.000	14:50
04 Std. Sol.	0.018	14:51
05 Room Air	0.000	14:51
06 Std. Sol.	0.020	14:52
07 Room Air	0.000	14:52

08 Sim Temp = 34.0°C

Simul Ser No = DR3378
Std Sol No = 18020
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Low AC*
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005941
Location = TOXL 8164.14.00 09/16
06/24/2019 14:53

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:54
02 Std. Sol.	0.300	14:54
03 Room Air	0.000	14:55
04 Std. Sol.	0.303	14:56
05 Room Air	0.000	14:56
06 Std. Sol.	0.301	14:57
07 Room Air	0.000	14:58

08 Sim Temp = 34.0°C

Simul Ser No = DR7350
Std Sol No = 201803H
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: HIGH AC
0.300 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005941
Location = TOXL 8164.14.00 09/16
06/24/2019 14:58

Test	AC	Time
01 Room Air	0.000	14:59
02 *Subject Test	INT*	15:00
03 Room Air	0.000	15:00

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS921456
Lot No = 34418080A2
Cyl No = 40
Expiration Date = 02/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Interferent Check*
CEE 0.000 0.10 AC + 0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005941
Location = TOXL 8164.14.00 09/16
06/24/2019 15:01

Test	AC	Time
01 Diagnostic	OK	15:02
02 Room Air	0.000	15:02
03 *Subject Test	RFI*	15:02
04 Room Air	0.000	15:03

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS921456
Lot No = 34418080A2
Cyl No = 40
Expiration Date = 02/05/2021
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005941
Location = TOXL 8164.14.00 09/16
06/24/2019 15:03

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:04
02 Std. Gas	0.081	15:04
03 Room Air	0.000	15:05
04 Std. Gas	0.082	15:05
05 Room Air	0.000	15:06
06 Std. Gas	0.081	15:06
07 Room Air	0.000	15:07

Lot No = 34418080A2
Cyl No = 40
Exp Date = 02/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005941
Location = TOXL 8164.14.00 09/16
06/24/2019 15:07

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:07
02 Std. Gas	0.081	15:08
03 Room Air	0.000	15:08
04 Std. Gas	0.081	15:09
05 Room Air	0.000	15:09
06 Std. Gas	0.081	15:09
07 Room Air	0.000	15:10

Lot No = 34418080A2
Cyl No = 40
Exp Date = 02/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005941
Location = TOXL 8164.14.00 09/16
06/24/2019 15:10

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:11
02 Std. Gas	0.081	15:11
03 Room Air	0.000	15:12
04 Std. Gas	0.081	15:12
05 Room Air	0.000	15:13
06 Std. Gas	0.081	15:13
07 Room Air	0.000	15:13

Lot No = 34418080A2
Cyl No = 40
Exp Date = 02/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION CHECK
0.080 AC

Form 106-I8000