

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-003057 Location: TOXL

A. General Setup and Checks:

- 1. Diagnostic Tests Pass and Instrument in Ready Mode
- 2. Breath tube heated
- 3. Date, time and location code (Level 2,E). Re-set if necessary.
- 4. Print test (Level 1,P). Sign and attach test record.
- 5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 417 psi Regulator: 400 psi

B. Tests (Sign and attach test records):

- Configure simulator for the following tests (Level 1,S).
 - 1. Wet Calibration Check
 - a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: DR3378 Lot #: 18020 Exp. Date: 1-9-20
 - b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.300 AC
Sim. Ser #: DR7350 Lot #: 20180314 Exp. Date: 3-22-20

Configure dry gas standard for the following tests (Level 1,S).

- 2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR3451 Lot #: ICS5 Exp. Date: NA
 Display reads "Interferent Detect"
- 3. RFI Check (CMS Mode) Display reads "RFI Detect"
- 4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 344180B0A2 Cyl #: 40 Exp. Date: 2.5.21
Test 1 0.079 Test 4 0.080 Test 7 0.079
Test 2 0.080 Test 5 0.079 Test 8 0.078
Test 3 0.079 Test 6 0.079 Test 9 0.079
Average 0.079

C. Remarks/Maintenance: ANNUAL INSPECTION

Instrument is acceptable to be used in the field.

Charles E. Ed
Breath Analyst Signature

6.25.2019
Date

NA

NA

Reviewed by

Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003057
Location = TOXL 8164.14.00 09/16
06/25/2019 13:19

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 34418080A2
Standard Cyl #? 40
Standard Expiration? 02/05/2021
Oper No? 666666

Flow Cal. Date: 04/04/2018
Slope 549
Intercept 197785

IR Calibration Date: 04/15/2015

	3um	9um
0th Coef(*100):	-23846	-15606
1st Coef(*100):	256102	135820
2nd Coef(*100):	2997	1528
H2O adj(mg/l*10k):	379	300

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003057
Location = TOXL 8164.14.00 09/16
06/25/2019 13:21

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:21
02 Std. Sol.	0.018	13:22
03 Room Air	0.000	13:23
04 Std. Sol.	0.018	13:23
05 Room Air	0.000	13:24
06 Std. Sol.	0.019	13:25
07 Room Air	0.000	13:25


08 Sim Temp = 34.0°C

Simul Ser No = DR3378

Std Sol No = 18020

County = 08

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

Low AC

0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003057
Location = TOXL 8164.14.00 09/16
06/25/2019 13:26

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:27
02 Std. Sol.	0.295	13:28
03 Room Air	0.000	13:28
04 Std. Sol.	0.296	13:29
05 Room Air	0.000	13:29
06 Std. Sol.	0.296	13:30
07 Room Air	0.000	13:31


08 Sim Temp = 34.0°C

Simul Ser No = DR7350

Std Sol No = 201830H

County = 08

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

HIGH AC
0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003057
Location = TOXL 8164.14.00 09/16
06/25/2019 13:33

Test	AC	Time
01 Room Air	0.000	13:34
02 *Subject Test	INT*	13:34
03 Room Air	0.000	13:35

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS921456
Lot No = 34418080A2
Cyl No = 40
Expiration Date = 02/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent Check
0.10AC + 0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003057
Location = TOXL 8164.14.00 09/16
06/25/2019 13:36

Test	AC	Time
01 Diagnostic	OK	13:37
02 Room Air	0.000	13:37
03 *Subject Test	RFI*	13:37
04 Room Air	0.000	13:38

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS921456
Lot No = 34418080A2
Cyl No = 40
Expiration Date = 02/05/2021
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003057
Location = TOXL 8164.14.00 09/16
06/25/2019 13:38

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:39
02 Std. Gas	0.079	13:39
03 Room Air	0.000	13:39
04 Std. Gas	0.080	13:40
05 Room Air	0.000	13:40
06 Std. Gas	0.079	13:41
07 Room Air	0.000	13:41

Lot No = 34418080A2
Cyl No = 40
Exp Date = 02/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003057
Location = TOXL 8164.14.00 09/16
06/25/2019 13:41

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:42
02 Std. Gas	0.080	13:42
03 Room Air	0.000	13:43
04 Std. Gas	0.079	13:43
05 Room Air	0.000	13:44
06 Std. Gas	0.079	13:44
07 Room Air	0.000	13:44

Lot No = 34418080A2
Cyl No = 40
Exp Date = 02/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003057
Location = TOXL 8164.14.00 09/16
06/25/2019 13:45

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:45
02 Std. Gas	0.079	13:46
03 Room Air	0.000	13:46
04 Std. Gas	0.078	13:47
05 Room Air	0.000	13:47
06 Std. Gas	0.079	13:47
07 Room Air	0.000	13:48

Lot No = 34418080A2
Cyl No = 40
Exp Date = 02/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000