





PROPERTY DISPOSAL
 OFFICE OF ATTORNEY GENERAL
 SFN 60567 (7/2016)

Multiple quantities of like items may be listed on one form and attach a sheet with tag numbers and serial numbers if available. Use a separate form for different items.

| | | |
|-------------------------------------|--|---|
| Division Crime Laboratory | Location of Property Breath Program | Date 7/5/19 |
| Point of Contact Charles E. Eder | Point of Contact Email ceder@nd.gov | Point of Contact Telephone Number 701-328-6159 |

| | |
|--|--|
| Description of Property (make, model, color, etc.) Intoxilyzer 8000 Breath Alcohol Instrument CMI, Inc., Owensboro, KY   | Tag Number (if available) <i>CE</i> Ag-3595 <i>AG 3505</i> |
| | Serial Number (if available) 80-005942 |
| | Date Purchased (if available) |
| | Original Cost (if available) <i>7394.00</i> |

Reason for disposal
 Issue with battery during power outage caused an issue with the power distribution board. Instrument no longer powers up.

Method of disposal:

- Landfill *(ELECTRONICS DISPOSAL)*
- Trash can/dumpster

| |
|--|
| Signature of person disposing property <i>Charles E. Eder</i> |
| Date 7/5/19 |

Send form to Michelle Metzger, Finance Division

| |
|-------------------------------------|
| Authorized Representative Signature |
| Date |