

ETHANOL GAS STANDARD CYLINDER REPORT

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM SEN 59282 (08/2018)

Chemical Test Operator Name (Print)	SRYAN N	AGEL	
Location 1201 12th STSC N	at-bidCt, ND	Intoxilyzer® Serial Nu	メローハハラグルト
Gas Lot Number 3441 8 0 8 0 A 2	Gas Cylinder Number	53	Gas Expiration Date 2-5-2021
Check When Done:	•		9 1
1. Scan/Enter Gas Cylinder Information	(Level 1, Function S).		
2. Perform an ACA Test (Level 1, Fund	ction C). (TEST		
Write in the ACA Test Results (Repo	itted to 3 Digits' Exitoro	1).	
B. 0. 0.75 AC			
C. 0.039 AC			
Complete the Top Portion of the Into	xilyzer® Record (SFN50	496, Form 120-G) and F	Place it by the intoxilyzer® for Use.
File Previous Intoxilyzer® Record (S	FN50496, Form 120-G)	at the Intoxilyzer® Locat	tion at the Agency.
5. Send the Following to the Crime Lab	oratory: (RECEIV	(D AT LAR)	
A. Ethanol Gas Standard Cylinder R	`		
7. 29.19 B. ACA Test Record.	_		
Chemical Test Operator Signature	-/-0		Date
Chemical Test Operator Signature	The X		7-29-2019
Reviewed By (Crime Laboratory Use Orly)	1	660	Date 7/22/12

Intoxilyzer Test Record and Checklist NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer North Dakota Model 8000 Location = WATF 07/29/2019

Alcohol Analyzer SN 80-005945 8164.14.00 09/16 03:36

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	03:36
02 Std. Gas	0.079	03:37
03 Room Air	0.000	03:37
04 Std. Gas	0.078	03:38
05 Room Air	0.000	03:38
06 Std. Gas	0.079	03:38
07 Room Air	0.000	03:39

Lot No = 34418080A2

Cyl No = 53

Exp Date = 02/05/2021

County = 27

Oper No. = 130626

Operator Signature BRYAN NAGEL

Remarks:

Form 106-I8000

STATE OF NORTH DAKOTA COUNTY OF MCKENZIE

i certify that the foregoing is a true and correct copy of the original document on file in this office. Dated at Wallord City, North Dakota

McKerzie County Sherin's Department