



# ETHANOL GAS STANDARD CYLINDER REPORT

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL

CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM

SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Trey N Nielsen</i>		
Location <i>PD3 PPD 1230 25th St S</i>	Intoxilyzer® Serial Number <i>80-006495</i>	
Gas Lot Number <i>34418080 A2</i>	Gas Cylinder Number <i>10</i>	Gas Expiration Date <i>2-5-2021</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).  
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).  
A. *0.082* AC  
B. *0.082* AC  
C. *0.082* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
  - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
  - B. ACA Test Record.

Chemical Test Operator Signature <i>Trey N Nielsen</i>	Date <i>10-14-19</i>
Reviewed By (Crime Laboratory Use Only) <i>Charles E. Ed</i>	Date <i>10-22-2019</i>

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006495  
Location = FAPD      8164.14.00 09/16  
10/14/2019      15:32

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:33
02 Std. Gas	0.082	15:33
03 Room Air	0.000	15:34
04 Std. Gas	0.082	15:34
05 Room Air	0.000	15:35
06 Std. Gas	0.082	15:35
07 Room Air	0.000	15:36

Lot No = 34418080A2  
Cyl No = 10  
Exp Date = 02/05/2021  
County = 09      Oper No. = 110104

  
Operator Signature  
TROY NIELSEN

Remarks:

Form 106-I8000